



Adherium (NZ) Limited
Level 2, 204 Quay Street, Auckland 1010, New Zealand
PO Box 106-612, Auckland 1143, New Zealand
Phone +64 9 307 2771
contact@smartinhaler.com
www.smartinhaler.com

12 March 2019

EU Declaration of Conformity

Radio Equipment Directive 2014/53/EU

Radio equipment and object of this declaration:

Product Name: **Hailie™ Sensor (for Bevespi Aerosphere®) / Bevespi+**
Product Type: **MDI Accessory**
Description: **Medication Usage Monitor & Reminder for Metered Dose Inhaler**
Model Number: **NF0096-Y-X-X**
(where x is any characters and Y is WH)

Manufacturer: Adherium (NZ) Ltd, Level 2, 204 Quay Street, Auckland 1010, New Zealand.

This declaration of conformity is issued under the sole responsibility of the manufacturer.

The object of the declaration described above is in conformity with the relevant Union harmonisation legislation: Directive 2014/53/EU.

The object of the declaration is in conformity with the following harmonised standards:

EN 300 328	Spectrum (RED Article 3.2)
EN 301 489-1	EMC (RED Article 3(b))
EN 301 489-17	EMC (RED Article 3(b))
EN 60601-1	Health & Safety (RED Article 3.1(a))

Chris Mander
Head of Regulatory and Quality
Adherium (NZ) Ltd