

## Regulatory Compliance Testing (OAR-333-007) Official Report

### 23mg CBD - FSB23TN23101 Duplicate

Sample ID: G3H0023-06 Test ID: 5028540 Source ID:

Date Sampled: 07/31/23

Date Accepted: 07/31/23

Matrix: Hemp Products

## Flora Sophia Botanicals, LLC

joannaf@florasophiabotanicals.com

$\sim$ $  -$		1		%RPD
	Averaged	Primary	Duplicate	(10% Action Level)
Total THC:	<loq(0.0040%) %<="" td=""><td>&lt; LOQ %</td><td>&lt; LOQ %</td><td>0 % PASS</td></loq(0.0040%)>	< LOQ %	< LOQ %	0 % PASS
Total CBD:	2.618 %	2.589 %	2.647 %	2.22 % PASS



Eric Wendt Chief Science Officer - 8/7/2023

Page 1 of 4

These results relate only to the sample included on this report. The report may not be reproduced except in full, without the written permission of Green Leaf Lab.



## Regulatory Compliance Testing (OAR-333-007) Official Report

### 23mg CBD - FSB23TN23101 Duplicate

Sample ID: G3H0023-06 Test ID: 5028540 Source ID:

Date Sampled: 07/31/23

Matrix: Hemp Products

Date Accepted: 07/31/23

## Flora Sophia Botanicals, LLC

joannaf@florasophiabotanicals.com

ate/Time Extra	cted: 08/02/		ency Analysis by HPLC Analysis Method/SOP: 215	Batch Identification: 2331037
Cannabinoids	LOQ (%)	mg/g	Cannat	pinoids Profile
Total THC	0.0040	< LOQ		
Total CBD	0.0040	26.47		
THCA	0.0040	< LOQ		
delta 9-THC	0.0040	< LOQ		
delta 8-THC	0.1534	< LOQ		
THCV	0.1197	< LOQ		
THCVA	0.1792	< LOQ		
CBD	0.0040	26.47		X
CBDA	0.0040	< LOQ		
CBDV	0.1232	< LOQ		CBD 2.6 Total: 2.6
CBDVA	0.1692	< LOQ		
CBN	0.1106	< LOQ		
CBG	0.1287	< LOQ		
CBGA	0.1705	< LOQ	2.6	
CBC	0.1614	< LOQ		
tal THC = delta tal CBD = CBD tal CBG = CBG -	+ (CBDA * 0.87	7)		

#### LOQ=Limit of Quantification, the lowest measurable concentration of an analyte.



Eric Wendt Chief Scien

Chief Science Officer - 8/7/2023



Regulatory Compliance Testing (OAR-333-007) Official Report

# Quality Control Potency

#### Batch: 2331037 - 215-Products

Blank(2331037-BLK1)							
Analyte	Result	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes
THCA	< LOQ	0.0039	%		08/02/23 12:01	08/02/23 22:29	
delta 9-THC	< LOQ	0.0039	%		08/02/23 12:01	08/02/23 22:29	
delta 8-THC	< LOQ	0.1510	%		08/02/23 12:01	08/02/23 22:29	
THCV	< LOQ	0.1178	%		08/02/23 12:01	08/02/23 22:29	
THCVA	< LOQ	0.1763	%		08/02/23 12:01	08/02/23 22:29	
CBD	< LOQ	0.0039	%		08/02/23 12:01	08/02/23 22:29	
CBDA	< LOQ	0.0039	%		08/02/23 12:01	08/02/23 22:29	
CBDV	< LOQ	0.1212	%		08/02/23 12:01	08/02/23 22:29	
CBDVA	< LOQ	0.1665	%		08/02/23 12:01	08/02/23 22:29	
CBN	< LOQ	0.1089	%		08/02/23 12:01	08/02/23 22:29	
CBG	< LOQ	0.1266	%		08/02/23 12:01	08/02/23 22:29	
CBGA	< LOQ	0.1679	%		08/02/23 12:01	08/02/23 22:29	
CBC	< LOQ	0.1589	%		08/02/23 12:01	08/02/23 22:29	
Poforonco(2331	027 CDM4)						

#### Reference(2331037-SRM1)

Analyte	% Recovery	LOQ	Units	%Recovery Limits	Extracted	Analyzed	Notes	
THCA	104	0.0098	%	90-110	08/02/23 12:01	08/02/23 22:52		
delta 9-THC	101	0.0098	%	90-110	08/02/23 12:01	08/02/23 22:52		
delta 8-THC	99.0	0.3773	%	90-110	08/02/23 12:01	08/02/23 22:52		
CBD	95.7	0.0098	%	90-110	08/02/23 12:01	08/02/23 22:52		
CBDA	90.7	0.0098	%	90-110	08/02/23 12:01	08/02/23 22:52		





Eric Wendt Chief Science Officer - 8/7/2023



## **Notes and Definitions**

Regulatory Compliance samples were collected onsite at facility according to ORELAP-SOP-001 and ORELAP-SOP-002 and following Sampling Plan FN117. Quality Control samples were tested as received. Results do not include uncertainty of measurements. Available upon request.

- ATM Non-cannabis matrix related interference or suppression of Internal standard
- BLI Baseline Interference Cannabinoid peak interference in chromatographic baseline affecting QC recovery .
- BLK Analyte detected in method blank, but not associated samples.
- BSH Blank Spike High Blank Spike recovery above method limit. no detections in samples.
- BSL Blank Spike Low Blank Spike recovery below lower method limit, analyte chromatography reviewed
- C manually for all samples.
- CBD Interference due to co-elution
- CV1 CBD matrix interference on GC Pest chromatography
- CV2 CCV was above acceptance criteria, Non-detect samples are considered acceptable.
- INF CCV was below acceptance criteria, sample still exceeds regulatory limit.
- ISH One or more QC falls outside acceptance criteria. Data entered into LIMS for informational purposes only.
- ISL Internal Standard concentration is above acceptance criteria.
- MSH Internal Standard concentration is below acceptance criteria.
- MSI Matrix Spike High Matrix Spike recovery above method limits.
- MSL Matrix Spike Interference Matrix spike source sample contains analyte hit above calibration affecting
- TPP recovery accuracy in Matrix Spike.
- U Matrix Spike Low Matrix Spike recovery below lower method limit, analyte chromatography reviewed manually for all samples.

Internal Standard concentration outside control limit due to matrix interference





Eric Wendt Chief Science Officer - 8/7/2023