

IDENTCERAM[®]

CERTIFICATE



DIVISION OF THE IDENTALLOY COUNCIL

The manufacturer certifies that the ceramic material provided to the laboratory with this certificate is FDA-listed Yttrium Zirconia.

The laboratory certifies that the specified manufacturer, materials, systems, and procedures were used to fabricate this prosthesis.

Material / Composition

Material / System Manufacturer

ArgenZ ST Multilayer	
$ZrO_2 + HfO_2 + Y_2O_3$ >99wt%	
Y_2O_3 8.5-10wt%	Al_2O_3 <0.1wt%
HfO_2 <5wt%	Fe_2O_3 <0.1wt%

DENTIST CERTIFICATE (attach to patient record)
www.identalloy.org

IDENTCERAM[®]

CERTIFICATE

DIVISION OF THE IDENTALLOY COUNCIL



Material / Composition

ArgenZ ST Multilayer

$ZrO_2 + HfO_2 + Y_2O_3$	>99wt%	
Y_2O_3	8.5-10wt%	Al_2O_3 <0.1wt%
HfO_2	<5wt%	Fe_2O_3 <0.1wt%

Material / System Manufacturer

The Argen Corporation

LAB COPY

73767 Rev.1 8/19