

# IDENTOCERAM<sup>®</sup>

## CERTIFICATE

DIVISION OF THE IDENTALLOY COUNCIL



The manufacturer certifies that the ceramic material provided to the laboratory with this certificate is FDA-listed Yttrium Zirconia.

The laboratory certifies that the specified manufacturer, materials, systems, and procedures were used to fabricate this prosthesis.

### Material / Composition

ArgenZ ST Multilayer

ZrO<sub>2</sub>+HfO<sub>2</sub>+Y<sub>2</sub>O<sub>3</sub> >99wt%  
Y<sub>2</sub>O<sub>3</sub> 8.5-10wt%  
HfO<sub>2</sub> <5wt%

Al<sub>2</sub>O<sub>3</sub> <0.1wt%  
Fe<sub>2</sub>O<sub>3</sub> <0.1wt%

### Material / System Manufacturer

The Argen Corporation

# IDENTOCERAM<sup>®</sup>

## CERTIFICATE

DIVISION OF THE IDENTALLOY COUNCIL



73767 Rev.1 8/19

### Material / Composition

ArgenZ ST Multilayer

ZrO<sub>2</sub>+HfO<sub>2</sub>+Y<sub>2</sub>O<sub>3</sub> >99wt%  
Y<sub>2</sub>O<sub>3</sub> 8.5-10wt%  
HfO<sub>2</sub> <5wt%

Al<sub>2</sub>O<sub>3</sub> <0.1wt%  
Fe<sub>2</sub>O<sub>3</sub> <0.1wt%

### Material / System Manufacturer

The Argen Corporation

LAB COPY

DENTIST CERTIFICATE (attach to patient record)

[www.identalloy.org](http://www.identalloy.org)