

IDENTCERAM CERTIFICATE

DIVISION OF THE IDENTALLOY COUNCIL



The manufacturer certifies that the ceramic material provided to the laboratory with this certificate is FDA-listed Yttrium Zirconia.

The laboratory certifies that the specified manufacturer, materials, systems, and procedures were used to fabricate this prosthesis.

Material / Composition

| | |
|--------------------------------|-----------|
| ZrO ₂ | 94%~95% |
| Y ₂ O ₃ | 4.5%~5.5% |
| Al ₂ O ₃ | <0.5% |
| Other Oxide | <0.5% |

Material / System Manufacturer

Haus Milling Center Inc.

DENTIST CERTIFICATE (attach to patient record)

www.identalloy.org

IDENTCERAM CERTIFICATE

DIVISION OF THE IDENTALLOY COUNCIL



Material / Composition

HAUS HT Plus Zirconia

| | |
|--------------------------------|-----------|
| ZrO ₂ | 94%~95% |
| Y ₂ O ₃ | 4.5%~5.5% |
| Al ₂ O ₃ | <0.5% |
| Other Oxide | <0.5% |

Material / System Manufacturer

Haus Milling Center Inc.

LAB COPY