

# IDENTALLOY<sup>®</sup>

## CERTIFICATE



The manufacturer certifies that the dental casting alloy provided to the laboratory with this certificate is a Predominantly Base alloy (contains less than 25% [total] Gold, Palladium and Platinum).

The laboratory certifies that Predominantly Base alloy was used to fabricate this prosthesis.

Alloy	Manufacturer	Composition %
-------	--------------	---------------

MoguCera C Disc	S&S Scheftner GmbH	Co 65% Si <1%
		Cr 28% Nb <1%
		Mo 5% Mn <1%
		C <1% Fe <1%

DENTIST CERTIFICATE (attach to patient record)  
[www.identalloy.org](http://www.identalloy.org)

# IDENTALLOY<sup>®</sup>

## CERTIFICATE



53863 Rev.1 2/16

### ALLOY

MoguCera C Disc

**MANUFACTURER**  
S&S Scheftner  
GmbH

### COMPOSITION %

Co	65%	Si	<1%
Cr	28%	Nb	<1%
Mo	5%	Mn	<1%
C	<1%	Fe	<1%

LAB COPY