
Application Information

Applicant Elitech Technology Inc.

Sample Information Digital pressure gauge, Sample Type: PGW-500/PGW-800

Test Project Pressure limit

Edit Date 2018.02.02

Remark The above information is provided and confirmed by the applicant.

Test Result

No.	Test Project	Result
	Digital pressure gauge (PGW-500)	
1	Pressure limit, MPa	≥3.6
	Digital pressure gauge (PGW-800)	
2	Pressure limit, MPa	≥6
Remark	/	

End of Form

Notes:

- 1.The research result is only responsible for the provided sample, not including DUT replication characteristics. The research result only represents the evaluation of tested sample, and can only be used for scientific research, data analysis, performance evaluation, quality analysis. We will not be responsible for any economical or legal liability generated from direct or indirect usage of the testing result.
- 2.The unrepeatable test shall not be retested. The applicant shall give up the right of dissent.
- 3.The research result does not have a social proof, if it is required in the process of trial or arbitration, the applicant shall inform us before proving samples. Otherwise, we will not be liable for all losses and disputes and reserve the right to claim for proper compensation.
- 4.Any unauthorized copy, illegal transfer, embezzlement, imposture, modification or any other kind of manipulation in full or part is not valid. We will reserve the right to the legal responsibility.

*****The End*****

