


Product Overview

Disclaimer! Please note, dependent on purchased model, some functionalities differentiate based on sensor configuration.

Thank you for selecting the **CM-500 handheld multi gas detector**. With long-term data storage, the CM-500 is designed to simultaneously measure multiple gas concentrations in the ambient environment. When the CM-500 measures gas concentrations reaching the alarm setting or higher, the **audible** alarm functions **will be** activated. The Handheld Multi Gas Detector can **also** be used **throughout** a wide range of applications, which include, **restaurant, beverage, breweries, agriculture, laboratory, construction, safety** and more.

Top Features:

- NDIR (CO2), EC (CO, NH3), and Fluorescence quenching by Oxygen (O2) is used to measure gas concentrations.
- Measurement items:** Basic: CO2, CO, PM2.5, PM10, RH, DP, AMB, Barometer
- Optional:** O2, NH3
- Large LCD for easy reading of gas concentration and temperature and humidity
- Audible alarm
- Data logging with SD card
- Built in the LCD back-light for easy reading in the dark
- Li- ion 18650 3.7V rechargeable batteries for longer lifespan



- A. Power Switch
- B. Humidity Sensor
- C. Temperature Sensor
- D. LCD Display
- E. Charge Indicator
- F. M1 (Zone 1 Mode Key)
- G. UP/TWA/STEL/MAX/MIN
- H. M2 (Zone 2 Mode Key)
- I. Enter Key
- J. DOWN/ALTI
- K. Log Key (Data Logger)
- L. USB Socket
- M. SD Card Slot
- N. RS485 Internal Jack
- O. Battery Cover
- P. Desktop Stand
- Q. Screw Position
- R. Tripod Screw

Key description:

M1 (Ⓜ)	Zone 1 sensor selection (CO ₂ , CO, O ₂ , AMB)
M2 (Ⓜ)	Zone 2 sensor selection (CO ₂ , CO, O ₂ , NH ₃ , RH, DP, AMB, PM2.5, PM10)
▲ (Ⓜ)	TWA (CO ₂ , 8-hr time weighted average), STELL ON DEVICE, (CO ₂ , 15 min. weighted average), MAX, MIN
▼ (Ⓜ)	ALTI (Atmospheric pressure)
Enter (Ⓜ)	Execute a command or select options.
Log (Ⓜ)	Data logger

LCD display symbol description:

Symbol	Description	Basic	Optional
	Buzzer On	<input type="checkbox"/>	<input type="checkbox"/>
CO ₂ 400 ppm	CO ₂ concentration, parts per million (ppm)	<input type="checkbox"/>	<input type="checkbox"/>
CO 80 ppm	CO concentration, parts per million (ppm)	<input type="checkbox"/>	<input type="checkbox"/>

PM2.5 18 ug/m ³	PM2.5 dust concentration (µg/m ³)	<input type="checkbox"/>
PM10 18 ug/m ³	PM10 dust concentration (µg/m ³)	<input type="checkbox"/>
O ₂ 20.08 %	O ₂ concentration (%)	<input type="checkbox"/>
NH ₃ 90 ppm	NH ₃ concentration ppm (parts per million)	<input type="checkbox"/>
	Battery capacity	<input type="checkbox"/>
ALTI BARO	Atmospheric pressure	<input type="checkbox"/>
TWA	Time weighted average (8 hours)	<input type="checkbox"/>
STELL	Short-Term Exposure Limit (15 min. weighted average)	<input type="checkbox"/>
MAX	Maximum value	<input type="checkbox"/>
MIN	Minimum value	<input type="checkbox"/>
AMB	Ambient temperature	<input type="checkbox"/>
RH	Relative humidity	<input type="checkbox"/>
DP	Dew point	<input type="checkbox"/>
TIME	Real-time setting display	<input type="checkbox"/>
YEAR	Year display	<input type="checkbox"/>
CAL L H	High/low gas concentration calibration	<input type="checkbox"/>
AL L H	High/low gas concentration alarm	<input type="checkbox"/>
	Display cycle	<input type="checkbox"/>
	SD card	<input type="checkbox"/>

Operation Instructions


Power on:
Switch the Power Key (A) on.

Measurement:
After power on, the device starts to measure and update the data every second.
Press the "M1" key (Ⓜ) to switch the Zone 1 sensor display. (CO₂→CO→O₂→AMB)
Press the "M2" key (Ⓜ) to switch the Zone 2 sensor display. (CO₂→CO→O₂→NH₃→RH→DP→AMB→PM2.5→PM10)

°C/°F switch:
When the temperature is displayed on the screen, press the "▲" key (Ⓜ) to switch °C/°F.

Backlight:
The backlight appears on the screen during startup and operation. After the device is idle for 30 seconds, the backlight is automatically turned off to extend battery life.


Display all sensor readings
Press the "M2" key (Ⓜ) for more than 3 seconds and the screen displays . Zone 2 displays readings of all sensors sequentially in time intervals. If you need to increase the display interval time, long press the "M2" key (Ⓜ). The interval time will automatically increase (1+0.5+0.5..... max. 10 seconds), and release the "M2" key (Ⓜ), the readings will be displayed according to the last set interval. If press the "M2" key (Ⓜ) again for more than 3 seconds, the function can be cancelled.



Real-time setting:
Long press the "M1" key (Ⓜ) to proceed the date and time setting. The default format is 24-hour. Press the "▲" (Ⓜ)/"▼" (Ⓜ) key to edit the parameters, and press the "Enter" key (Ⓜ) to proceed the next setting. The setting sequence is: year→month→day→hour→minute. After finishing the real-time setting, press the "M1" key (Ⓜ) to continue the alarm setting.




Alarm settings:
The setting order of the sensors is CO₂→CO→O₂→NH₃→PM2.5. Press the "Enter" key (Ⓜ) to switch the high/low gas concentration alarm setting. Press "▲" (Ⓜ)/"▼" (Ⓜ) key to edit the parameters, and then press the "M1" key (Ⓜ) to proceed the next sensor alarm value settings. After settings, press and hold the "M1" key (Ⓜ) to return to the main screen.



Data logging:
Press the "Log" key (Ⓜ) for 3 seconds and the screen displays to start the data logging function, long press the "Log" key (Ⓜ), this function can be canceled. When flashes, it means that the SD card is broken or full; when remains stationary, it indicates normal operation.



Data logging interval time setting:
The default interval time is 00:02 (2 seconds). Press the "Log" key (Ⓜ) for more than 3 seconds and then the interval time will automatically increase. The interval time, loop display, ranges from 00:02 seconds to 60:00 minutes and then back to 00:02 seconds. The interval time ≤5 minutes, the buzzer on/off switch is controlled by the "Enter" key (Ⓜ): the interval time is ≥5 minutes, the buzzer is off.



Atmospheric pressure display:
Press "▼" key (Ⓜ), to show barometric pressure.
Atmospheric pressure compensation:
Press and hold the "▼" key (Ⓜ) to turn the compensation function on/off. The gas compensation is only for CO₂ and O₂.

Buzzer on/off:
Press the "Enter" key (Ⓜ) to turn the buzzer on/off. During operation, if AL L or AL H buzzer sounds, stops to operate. At this time, press the "Enter" key (Ⓜ) to turn off the buzzer and continue . In addition, the buzzer will be automatically turned off when the data logging interval time is ≥5 minutes.




Buzzer AL L/AL H:


When AL L is on, the buzzer sounds "Ta Ta Ta"; when AL H is on, the buzzer sounds "Bi Bi Bi".

RS485 **INTERNAL** communications:

Please use the 4 Pole 3.5mm headphone plug. The headphone plug contacts are defined as below:



For long-distance communication, it is necessary to use a dedicated isolated line to ensure communication quality. The wiring method is as below:



Power off:


Switch the Power Key (A) off.

Note: When the device is charged with USB, it cannot be turned off.

Calibration:


(1) Single-point calibration

At the same time, long press "▲ (G)" and "▼ (J)" key, CAL is displayed on the screen. Press the "M1" key (E) to select the sensor that needs to be calibrated. Press the "▲ (G)/▼ (J)" key to edit the parameters, and then press the "Enter" key (I) to proceed the next setting. Zone1 displays the unadjusted raw value; Zone2 displays the adjusted value, and it is also the value of returning to the main screen; Zone1-Zone2=adjustment amount. After finishing the setting, press and hold the "M1" key (E) to return to the main screen.



(2) Two-point calibration


At the same time, long press "▲ (G)" and "▼ (J)" key, CAL is displayed on the screen. And then long press the "M2" key (H), CAL L is displayed on the screen. Next, press the "M1" key (E) to select the sensor that needs to be calibrated (CO2→CO→O2→NH3→PM2.5→RH→AMB) and the corresponding unit will blink. Press "▲ (G)/▼ (J)" key to adjust the value of the standard gas concentration. Press "Enter" key (I) to switch CAL H (High gas concentration calibration)/CAL L (low gas concentration calibration). The ppm unit is not flashing for a moment, indicating that the calibration value has been stored. Long press the "M2" key (H) for a calibration calculation. At the same time, the screen displays the calibration reading to confirm the accuracy. Short press the "M1" key (E) to set the next gas calibration, or long press the "M1" key (E) to return to the main screen.



Clear Two-point calibration:

Set the same calibration value for CAL L and CAL H, and then proceed the calculation.

After the calibration is completed, the original calibration value is cleared.



Rechargeable Batteries

Battery message:



'Battery OK':

Measurements are possible



'Battery Low':

The battery needs to be recharged, measurements are still possible



'Battery Exhausted':

Measurements are not possible

Battery installation:

This device is supplied with Li-ion 18650 3.7V rechargeable battery *3 pcs.

Please confirm whether batteries + polarity are Li ion18650 positive bump specification and install batteries in the correct polarity. If the battery polarity is reversed, the capacity will be insufficient and the battery life will be shortened.

Battery charging:

The charge lamp (E) lights up while charging; the charge lamp (E) goes off when charging is completed.

(5V/1A USB adapter charger)

Note: During battery charging, the temperature of the device will rise by 5°C-10°C. At this time, the measurements of temperature and humidity will be affected by temperature rise. Cause an impact on the accuracy of temperature when charging, please use a fan to blow toward the temperature sensor (C) directly in order to get a compensated balance of temperature and humidity between temperature sensor and surrounding area.

Safety Instructions

Warning: Your safety is very important to us. To ensure to use the device correctly and safely, we would like to draw your attention to read the warning and entire User Manual before using the device. These are important safety information and should be observed at all times.

1. Please handle the device lightly, do not subject the device to impact or shock.
2. Do not immerse the device in water. Water can cause electric shock, fire or malfunction which may result in damage.
3. Do not keep the device under the hot and moisture environment. Keep the device away from the heat source or near water.
4. Please use a standard USB power supply (such as PC's USB port, universal AC adapter with USB port). Improper power supply can cause serious damage to the device, or result in injury or death to the user.

Product Care

To ensure you receive the maximum benefit from using this device, please observe the follow guidelines.

1. The maximum capacity of the SD card is **16G**.
2. During battery charging, the temperature of the device rises. The temperature and humidity sensors **will be** affected. At this time, measurements are only for reference. After batteries are fully charged and the device cools down, measurements are **again**, reliable.
3. Repair – Do not attempt to repair the device or modify the circuitry by yourself. Please contact with the local **manufacturer for repairs - Support@CO2Meter.com**
4. Caution – The CO and NH3 sensors must be replaced **every** 3 years.
5. Cleaning – Disconnect the power before cleaning. Use a damp cloth, do not use the liquid cleaning agent, such as **benzene, thinner or aerosols**.
6. Maintenance – Recommend that the user conducts a comprehensive test and calibration every year to **ensure normal operation of the device**.

Specifications

	Measurement Range	Accuracy	Display Resolution
CO ₂	0-9,999ppm (5,001-9,999ppm over range)	±50ppm or 5% of reading, whichever is greater (0-5,000ppm)	1ppm
CO	0-1,000 ppm	±5%-±10%	1ppm
O ₂	0-25%	<2% FS/0.1 mbar	0.01%
NH ₃	0-100ppm	±10%	1ppm
PM2.5 PM10	999 ug/m ³	±15% or ±15 ug/m ³ , whichever is greater	1 ug/m ³
RH	0-100% RH	±3%@25°C (20-80% RH), others ±5%	0.01%
Temperature	0-50°C	±1°C	0.01°C
Barometer	50-110 kPa	±0.4kPa	0.1 mmHg
Operating Conditions	0-40°C (32-104°F), 0-95% RH, non-condensing, with lithium-ion batteries 0-50°C (32-122°F), 0-95% RH, non-condensing, without lithium-ion batteries		
Storage Temp.	-20-60°C (-4-140°F), 0-85% RH, non-condensing, without lithium-ion. batteries		
Power Supply	Rechargeable Battery : Li-ion 18650 3.7V (device uses 3) AC Adapter : 5V±5% >1A, 100-240 VAC, 50/60 Hz		
Storage Capacity	depend on SD card capacity (max.16G SD card)		
Comm. Interface	RS485 ModBus BR19200 · N · 8 · 1 (INTERNAL)		
Alarm	80db±5%@10cm		

Volume	
Weight	305g (without batteries)
Dimension	204.5x91.7x49.6 mm (8.1x3.6x2 inch)

Note: After power-on, it would take 20 minutes for the device to stably measure the temperature and RH.

EMC/RFI: Readings may be affected if the unit is operated within radio frequency electromagnetic field strength of approximately 3 volts per meter, but the performance of the instrument will not be permanently affected.

Sensor warm-up time and logging Interval:

No.	Sensor	Warm-up time	Data logging interval time setting	
			After powering on and standby time, data logging can be operated.	Data logging interval time setting
1	CO ₂	< 1 minute	After power on 5 minutes	2 seconds-60 minutes
2	CO	5 minutes	After power on 5 minutes	2 seconds-60 minutes
3	O ₂	< 1 minute	After power on 5 minutes	2 seconds-60 minutes
4	NH ₃	5 minutes	After power on 5 minutes	2 seconds-60 minutes
5	PM2.5 PM10	< 1 minute	After power on 5 minutes	2 seconds-60 minutes

Installation Instructions

You can place the CM-500 series on a table (please pull out the Housing Stand (P) or hang it on the wall.

Please note the following: When hanging the CM-500 on the wall, it is important to choose a suitable location to install the device, next simply fix a screw to the wall and begin monitoring.

Package Contents

- ▶▶ Handheld Multi Gas Detector
- ▶▶ Manual
- ▶▶ USB 1.5M cable
- ▶▶ SD Card

Support

The quickest way to obtain technical support is via email. Please send all support inquiries to Support@CO2Meter.com.

In your email, please include a clear, concise, definition of the problem and any relevant troubleshooting information or steps taken so far so we can duplicate the problem and quickly respond to your inquiry.

For additional CM-500 Multi Gas Detector Information, you can reference below.

See CO2Meter, Inc. Terms & Conditions at, <https://www.co2meter.com/pages/terms-conditions>



CO2 Meter, Inc.
131 Business Center Drive,
Ormond Beach, FL 32174
Sales@CO2Meter.com | www.CO2Meter.com



Ref.No. : 122018
07-21-20