

Quality Control Certificate

| Product: | STERILE PLASTIC FORCEP | | |
|----------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------------|--------------------------------------|-------|------------|
| Component Number: | B21809-ETO | | |
| Batch / Lot Number: | 74382 | | |
| ETO Batch No: | WA066747 | | |
| Reduction Factor: | 1675 | | |
| | | | |
| Has passed QC batch testing as specified by our Group QA Acceptance outlined below (tick as appropriate): | | | |
| BS 6002-1:2007 "Sampling procedures for inspection by variables" for Batch WA066747 | | | |
| Low DNA template Human DNA contamination at 31 cycles amplification. | | | |
| Product is compliant to the requirements of ISO 18385:2016* and this batch has been assessed to meet the DNA reduction requirements of this standard by an ISO 17025 accredited forensic (Key Forensics) analytical testing laboratory." | | | |
| Signature: | Robert | Date: | 04/02/2021 |
| Print Name: | Kayleigh Probert | | |
| Position: | QC Standards & Compliance Department | | |
| *In an audit led by Transforming Forensics of the production of ETO-treated DNA consumables at the SceneSafe facilities at Burnham on Crouch, in March 2020 the processes audited were assessed to be compliant with the quality standard ISO18385:2016. | | | |



This is to certify that

