



PRODUCTS (UK) LTD ... Evidently Better



# Quality Control Certificate

This is to certify that

Product: Spiked Swabs  
 Component Number: B22711 KFS/19/53025/01  
 Batch / Lot Number: 19C04  
 ETO Batch No: WA057648

Has passed QC batch testing criteria as specified by our Group QA Acceptance Criteria outlined below (tick as appropriate):

- BS 6002-1:2007 "Sampling procedures for inspection by variables" for Batch WA057648
- Low DNA template Human DNA contamination at 30 cycles amplification.
- Certified product is confirmed as being ISO 18385:2016 Forensic DNA Grade and has been tested to this standard by an ISO 17025 accredited forensic analytical testing laboratory.

QC Test Signature:                     M. Elwell                     Date:                     10/7/19                    

Print Name: Marion Elwell / ~~Terry Bell~~ (delete as appropriate)

Position: QC Standards & Compliance Department

Authorisation:                     John Carpenter                     Date:                     10/7/19                    

Print Name: John Carpenter

Position: WA Group Operations Director

WA GROUP  
 MIDAS HOUSE  
 8 & 9 BURNHAM BUSINESS PARK  
 SPRINGFIELD ROAD  
 BURNHAM ON CROUCH  
 ESSEX, CM0 8TE

ENQUIRIES  
 TELEPHONE: 00 44 (0) 1621 786654  
 FACSIMILE: 00 44 (0) 1621 786655

