

... Evidently Better



Quality Control Certificate

This is to certify that			
Product:	Sterile Proctoscope Large		
Component Number:	B22508-ETO		
Batch / Lot Number:	WO65494/2018-19G01		
ETO Batch No:	WA058876		
Has passed QC batch testing criteria as specified by our Group QA Acceptance Criteria outlined below (tick as appropriate):			
BS 6002-1:2007 "Sampling procedures for inspection by variables" for Batch WA058876			
Low DNA template Human DNA contamination at 30 cycles amplification.			
Certified product is confirmed as being Forensic DNA Grade and has been tested to this standard by an ISO 17025 accredited forensic analytical testing laboratory.			
QC Test Signature:	M. Ewell	Date:	24/09/19
Print Name:	Marion Elwell / Terry Ball (delete as appropriate)		
Position:	QC Standards & Compliance Department		on lool to
Authorisation	John Capoule	Date:	24/09/19
Print Name:	John Carpenter		
Position:	WA Group Operations Director		
		SPRING BURNH	
			RIES ONE: 00 44 (0) 1621 7866 ILE: 00 44 (0) 1621 7866