

# Quality Control Certificate

This is to certify that

Product: Medical Wire MW102 Swab  
Component Number: B22711  
Batch / Lot Number: 20F08  
ETO Batch No: WA64310  
Reduction Factor: 3350

Has passed QC batch testing as specified by our Group QA Acceptance outlined below (tick as appropriate):

BS 6002-1:2007 "Sampling procedures for inspection by variables" for Batch WA64011

Low DNA template Human DNA contamination at 31 cycles amplification.

Product is compliant to the requirements of ISO 18385:2016\*   
and this batch has been assessed to meet the DNA reduction requirements of this standard by an ISO 17025 accredited forensic (Key Forensics) analytical testing laboratory."

Signature: M. Elwell Date: 13/8/2020

Print Name: Marion Elwell / ~~Kayleigh Probert~~ (delete as appropriate)

Position: QC Standards & Compliance Department

\*In an audit led by Transforming Forensics of the production of ETO-treated DNA consumables at the SceneSafe facilities at Burnham on Crouch, in March 2020 the processes audited were assessed to be compliant with the quality standard ISO18385:2016.

