

Quality Control Certificate

This is to certify that		
Product:	Condom Collection Kit	
Component Number:	G91613-O	
Batch / Lot Number:	71229	
ETO Batch No:	WA64310	
Reduction Factor:	3350	
Has passed QC batch testing as specified by our Group QA Acceptance outlined below (tick as appropriate):		
BS 6002-1:2007 "Sampling procedures for inspection by variables" for Batch WA64011		
Low DNA template Human DNA contamination at 31 cycles amplification.		
Product is compliant to the requirements of ISO 18385:2016* and this batch has been assessed to meet the DNA reduction requirements of this standard by a ISO 17025 accredited forensic (Key Forensics) analytical testing laboratory."		
Signature:	M. Ewell Date: 13/6/201	<u>'</u> <u>C</u>
Print Name:	Marion Elwell / Kayleigh Probert (delete as appropriate)	

*In an audit led by Transforming Forensics of the production of ETO-treated DNA consumables at the SceneSafe facilities at Burnham on Crouch, in March 2020 the processes audited were assessed to be compliant with the quality standard ISO18385:2016.

QC Standards & Compliance Department



Print Name:

Position:

