



广州市哈雷日用品有限公司  
Guangzhou Harley Daily Necessities Co., Ltd.

# 公司资料 (Company information)



## 营业执照

(副本)

编号: S2212019051300G(1-1)

统一社会信用代码  
91440114076502373X



扫描二维码登录  
“国家企业信用  
信息公示系统”  
了解更多登记、  
备案、许可、监  
管信息。

名称	广州市哈雷日用品有限公司	注册资本	叁佰万元 (人民币)
类型	有限责任公司(自然人投资或控股)	成立日期	2013年08月15日
法定代表人	徐美英	营业期限	2013年08月15日至 长期
经营范围	其他制造业 (具体经营项目请登录广州市商事主体信息公示平台查询, 网址: <a href="http://cri.gz.gov.cn/">http://cri.gz.gov.cn/</a> 。依法须经批准的项目, 经相关部门批准后方可开展经营活动。)		
		住所	广州市从化区太平镇元洲岗村 (厂房A01)

登记机关



2019年 03月 21日

国家企业信用信息公示系统网址: <http://www.gsxt.gov.cn>

市场主体应当于每年1月1日至6月30日通过国家企业信用信息公示系统报送公示年度报告

国家市场监督管理总局监制

528361

### 对外贸易经营者备案登记表

备案登记表编号: 03629647      进出口企业代码: 4401076502373

经营者中文名称	广州市哈雷日用品有限公司		
经营者英文名称	Guangzhou Harley Commodity Company Limited		
组织机构代码	076502373	经营者类型 (由备案登记机关填写)	私营有限责任公司
住所	广州市从化区太平镇元洲岗村 (厂房A01)		
经营场所 (中文)	广州市从化区太平镇元洲岗村 (厂房A01)		
经营场所 (英文)	(PLANT A01) YUANZHOU GANG VILLAGE, TAIPING TOWN, CONGHUA DISTRICT, GUANGZHOU		
联系电话	020-37512951	联系传真	020-37512951
邮政编码	510900	电子邮箱	139036088@qq.com
工商登记注册日期	2013-8-15	工商登记注册号	91440114076502373X

依法办理工商登记的企业还须填写以下内容


企业法定代表人姓名	徐美英	有效证件号	342423197511104505
注册资金	叁佰万元		(折美元)

依法办理工商登记的外国(地区)企业或个体工商户(独资经营者)还须填写以下内容

企业法定代表人/个体工商户负责人姓名		有效证件号	
企业资产/个人财产			(折美元)

备注

填表前请认真阅读背面的条款, 并由企业法定代表人或个体工商户负责人签字、盖章。



2019年 04月 10日

# ISO 9001:2015 certificate:

## 认证证书

标准 **ISO 9001:2015**

证书登记号码 **01 100 1430200**

证书持有者: **广州市哈雷日用品有限公司**  
 统一社会信用代码: 91440114076502373X  
 注册地址: 中国广东省广州市从化区太平镇元洲岗村 (厂房 A01)  
 邮编: 510900  
 经营地址: 同上述地址

认证范围: **口罩的设计和制造**

证明完成了审核并满足了 ISO 9001:2015 标准的要求。

有效期: 证书有效期从 2019-09-08 至 2020-08-31。  
 此证书须经过符合要求的监督审核保持有效。  
 初次发证始于 2014 年

本证书信息可在国家认证认可监督管理委员会官方网站上查询  
<http://www.cnca.gov.cn>

2019-09-10

*E. J. Huey*

TÜV Rheinland Cert GmbH  
 Am Grauen Stein · 51105 Köln

www.tuv.com



扫描全能王 创建



TÜV, TÜEV and TUV are registered trademarks. Utilisation and application requires prior approval.

# Certificate

Standard **ISO 9001:2015**

Certificate Registr. No. **01 100 1430200**

Certificate Holder: **Guangzhou Harley Commodity Company Limited**  
 Unified Social Credit Code: 91440114076502373X  
 Registration Address: (Plant A01) Yuanzhougang Village,  
 Taiping Town, Conghua District, Guangzhou City,  
 Guangdong Province 510900, P. R. China  
 Operation Address: same as above

Scope: **Design and Manufacturing of Respirator**

Proof has been furnished by means of an audit that the requirements of ISO 9001:2015 are met.

Validity: The certificate is valid from 2019-09-08 until 2020-08-31.  
 It remains valid subject to satisfactory surveillance audits.  
 First certification 2014

This certificate information can be searched on CNCA official website <http://www.cnca.gov.cn>

2019-09-10

*E. J. Huey*

TÜV Rheinland Cert GmbH  
 Am Grauen Stein · 51105 Köln

www.tuv.com



扫描全能王 创建

Environmental management system certificate:



### ENVIRONMENTAL MANAGEMENT SYSTEM CERTIFICATE

Registration No. 0350319E20131R1S

This is to certify that the environmental management system of **GUANGZHOU HARLEY COMMODITY CO., LTD.**

Taiping Town, Yuanzhougang Village, Shunhua District, Guangzhou City (Factory Building A01), 510900  
Social Credit Code: 91440114076502373X

is in conformity with  
GB/T 24001-2016 / ISO 14001:2015 Standard

This certificate is valid to  
Design and manufacturer of respirator (Site Covered: Taiping Town, Yuanzhougang Village, Shunhua District, Guangzhou City (Factory Building A01)).

Date of issue: June 06, 2019  
Date of expiry at most: April 26, 2022

Representative: *Wang Honglin*

XINGYUAN CERTIFICATION CENTRE CO., LTD. (XQCC)  
FOR CERTIFICATE  
(7FL, Tower C, Lihua Plaza, No. 9 Shangdi 3 St., Haidian, Beijing)



中国认可  
国际互认  
管理体系  
MANAGEMENT SYSTEM  
CNAS C035-M

NOTE: This certificate shall be maintained by regular surveillance audit. The validity of the certificate can be verified by scanning QR code. The information of the certificate can be available in <http://www.cnca.gov.cn>, the website of CNCA, and in our website <http://www.xqcc.com.cn>.



### 环境管理体系认证证书

注册号: 0350319E20131R1S

兹 证 明  
**广州市哈雷日用品有限公司**  
广州市从化区太平镇元洲岗村 (厂房A01), 510900  
统一社会信用代码: 91440114076502373X

管理体系符合  
GB/T 24001-2016 / ISO 14001:2015 标准

该管理体系适用于  
口罩的设计和制造\*\*\*

[ 涉及场所: 广州市从化区太平镇元洲岗村 (厂房A01) ]

颁证日期: 2019年6月6日  
有效期最长可至: 2022年4月26日

签发人: *王宏林*

兴原认证中心有限公司  
证书专用章  
(北京市海淀区上林三街1号新华大厦C座7层)



中国认可  
国际互认  
管理体系  
MANAGEMENT SYSTEM  
CNAS C035-M

注: 在证书有效期内, 获证组织须按规定接受年度监督审核, 保持认证资格。通过扫描二维码可获知证书的有效状态。该证书信息还可在国家认证认可监督管理委员会官方网站 ([www.cnca.gov.cn](http://www.cnca.gov.cn)) 和兴原认证中心有限公司官方网站 ([www.xqcc.com.cn](http://www.xqcc.com.cn)) 上查询。





产品介绍

Product introduction

N95 Respirator

**L-188**

N95 L188 mask picture:



# NIOSH N95 (L-188) certificate:TN-19824



DEPARTMENT OF HEALTH & HUMAN SERVICES

NIOSH Reference: TN-19824  
Mfr. Reference: GHCL-188

Centers for Disease Control  
and Prevention (CDC)  
National Institute for Occupational  
Safety and Health (NIOSH)  
National Personal Protective  
Technology Laboratory (NPPTL)  
P.O. Box 18070  
Pittsburgh, PA 15236-0070  
Phone: 412-386-4000  
Fax: 412-386-4051  
December 15, 2014

Mr. Wanghua Xu  
Management Representative  
Guangzhou Harley Commodity Company Limited  
Floor 1 to 4, #8, Team 17, Sandong Village  
Xinhua Street, Huadu District  
Guangzhou 510800  
CHINA

Dear Mr. Xu:

The National Institute for Occupational Safety and Health (NIOSH) has reviewed your request accepted September 17, 2014. This request was for approval of the model L-188 N95 filtering facepiece air purifying respirator for protections against particulates at a N95 filter efficiency level, reference assembly matrix L-188 Ama.xls. In addition, the Guangzhou Harley Commodity Company QA Manual, version A1, effective date 2014-06-10, was submitted for review.

This request is granted. Approvals are granted for English language only on all documentation. It is the manufacturer's responsibility to correctly translate materials desired in languages other than English. Approval number TC-84A-6973 has been assigned. This respirator is approved for protection against particulates at a N95 filter efficiency level.

NIOSH has also reviewed the quality manual presented and finds that this manual meets or exceeds the minimum technical requirements for quality assurance plans as outlined in Title 42, *Code of Federal Regulations* (CFR), Part 84.41(a) and, on the basis of this review, this quality manual is accepted.

The CD enclosed with this letter contains the final respirator label. The abbreviated label is accepted as submitted. The cautions and limitations which apply to this approval are on the approval label. Only those assemblies affected by this request, or where new approval numbers are assigned, apply to this approval action. Production approval labels cannot include information on unapproved configurations.

The approved assembly consists of the parts as listed on the approval label and the assembly matrix. Parts are to be marked with the numbers indicated on the approval label in a legible and permanent manner (marking cannot be removed without evidence of its previous presence).

Page 2- Mr. Wanghua Xu - TN-19824

This certificate of approval is not an endorsement of the respirator by NIOSH, and such endorsement shall not be stated or implied in advertisements or other publicity. However, you may publicize the fact that this respirator has met the requirements of Title 42, *Code of Federal Regulations*, Part 84 (42 CFR 84).

No changes may be made to any respirators and accompanying documentation without prior written approval of NIOSH. Requests for changes must be submitted to NIOSH and a modification of this approval must be granted before changes are made.

A copy of the quality manual will be retained by NIOSH and incorporated into our files. Any future changes to this accepted quality document must be submitted to NIOSH for a modification of this accepted quality system.

Sincerely yours,

Jonathan V. Szalajda  
Acting Chief, Technology Evaluation Branch  
National Personal Protective Technology Laboratory

Enclosures



# NIOSH N95 (L-188) test report:TN-19824



## TEST REPORT

**Task Number:** TN-19824

**Manufacturer:** Guangzhou Harley Commodity Company Limited

**Prepared by:** Nichole Petitta

**Tests Conducted by:** Nichole Petitta

**Date:** November 13, 2014

**Respirator Tested:** L-188 N95

### Background Information

Respirator tested as per test request.

### Tests Assigned

<u>Test Description</u>	<u>STP Number</u>
A. Exhalation Resistance Test	RCT-APR-STP-0003
B. Inhalation Resistance Test	TEB-APR-STP-0007
C. Sodium Chloride (NaCl) N95 Test	TEB-APR-STP-0059

### Overall Results

The respirator system tested met the requirements of all the procedures listed above.

**Task Number:** TN-19824

**Reference No.:** CFR 84.181

**Test:** Sodium Chloride (NaCl) - N95

**STP No.:** 59

**Manufacturer:** Guangzhou Harley Commodity Company Limited

**Item Tested:** L-188 N95

### **Comments:**

All samoles were tested on TSI 8130 machine 000192 with samples 1-3 using timer 000214.

Was all equipment verified to be in calibration throughout all testing?  Yes  No

Signature:

*Nichole L. Petitta*

Date: 11/13/2014

Engineering Technician

# NIOSH N95 (L-188) test report:TN-19824)

National Institute for Occupational Safety and Health  
Respirator Branch  
Test Data Sheet



Task Number: TN-19824 Reference No.: CFR 84.180  
Test: Exhalation Resistance Test STP No.: 3  
Manufacturer: Guangzhou Harley Commodity Company Limited

Filter Type: Filter Only

Item Tested: L-188 N95

Sample	Maximum Allowable Resistance (MM of H2O)	Actual Resistance (MM of H2O)	Result
	Exhalation	Exhalation	
1	25	9.6	PASS
2	25	8.9	PASS
3	25	8.5	PASS

Overall Result: PASS

Comments:

Samples were tested on Brooks meter 0000283.

Was all equipment verified to be in calibration throughout all testing?  Yes  No

Signature:

*Nicole L. Petitta*

Engineering Technician

Date: 11/4/2014

National Institute for Occupational Safety and Health  
Respirator Branch  
Test Data Sheet



Task Number: TN-19824 Reference No.: CFR 84.180  
Test: Inhalation Resistance Test STP No.: 7  
Manufacturer: Guangzhou Harley Commodity Company Limited  
Item Tested: L-188 N95

Filter Type: Filter Only

Sample	Maximum Allowable Resistance (MM of H2O)	Actual Resistance (MM of H2O)	Result
	Inhalation	Inhalation	
1	35	10.5	PASS
2	35	10.5	PASS
3	35	9.7	PASS

Overall Result: PASS

Signature:

*Nicole L. Petitta*

Engineering Technician

Date: 11/4/2014

# NIOSH N95 (L-188) test report:TN-19824

National Institute for Occupational Safety and Health  
Respirator Branch  
Test Data Sheet



Task Number: TN-19824 Reference No.: CFR 84.181  
Test: Sodium Chloride (NaCl) - N95 STP No.: 59  
Manufacturer: Guangzhou Harley Commodity Company Limited  
Item Tested: L-188 N95

Filter	Flow Rate	Initial Filter Resistance	Maximum Allowable Percent Leakage	Initial Percent Leakage	Maximum Percent Leakage	Result
1	85	9.6	5.00	0.517	0.517	PASS
2	85	10.6	5.00	0.694	0.694	PASS
3	85	9.8	5.00	0.718	0.718	PASS
4	85	10.3	5.00	0.651	0.654	PASS
5	85	9.3	5.00	0.682	0.682	PASS
6	85	9.7	5.00	0.755	0.755	PASS
7	85	10.7	5.00	0.621	0.621	PASS
8	85	9.9	5.00	0.744	0.744	PASS
9	85	9.8	5.00	0.535	0.535	PASS
10	85	10.8	5.00	1.500	1.580	PASS
11	85	10.5	5.00	0.688	0.688	PASS
12	85	9.7	5.00	0.692	0.692	PASS
13	85	10.6	5.00	0.757	0.757	PASS
14	85	10.8	5.00	0.671	0.671	PASS
15	85	10.8	5.00	0.981	0.981	PASS
16	85	10.3	5.00	0.678	0.678	PASS
17	85	9.8	5.00	0.661	0.661	PASS
18	85	10.3	5.00	0.680	0.680	PASS
19	85	9.6	5.00	1.660	1.660	PASS
20	85	9.8	5.00	0.769	0.769	PASS

Overall Result: PASS

Signature: *Siende L. Petitta* Date: 11/13/2014  
Engineering Technician

Task Number: TN-19824 Reference No.: CFR 84.180  
Test: Inhalation Resistance Test STP No.: 7  
Manufacturer: Guangzhou Harley Commodity Company Limited  
Item Tested: L-188 N95

Comments:  
Samples were tested on Brooks meter 000036.




Was all equipment verified to be in calibration throughout all testing?  Yes  No

Signature: *Siende L. Petitta* Date: 11/4/2014  
Engineering Technician

# L-188 实物图片 Pictures




# L-188 包装盒图片 Packing Pictures



盒规 (Box) : 20Pcs/Box, Size: 200\*145\*135mm

# L-188 合格证、箱规 Certificate & CTN Details



箱规 (CTN) : 20Boxes/CTN Size: 415\*415\*310=0.54m<sup>3</sup> G.W: 4.9KG