

Device Description

The physiology


The body's circulatory system serves to transport and distribute essential substances to the tissues of the body and to remove by-products of metabolism. It also plays a role in the regulation of body temperature, humoral communication throughout the body and adjustments of oxygen and nutrient supply in differing physiological states. The cardiovascular system is made up of a pump (the heart), a series of distributing and collecting tubes and an extensive system of thin vessels that allow rapid exchange with tissues. An average adult has a blood volume of about 5-6 litres. The venous system has a large capacity and may contain some 70% of the blood volume at any time with a large percentage of this in the lower legs. Cardiac output is the volume of blood pumped by the heart per minute and venous return is the volume returning to the heart in the same unit of time. These are interdependent and multiple feedback control loops operate to regulate the cardiovascular system. Ancillary factors can affect venous return including muscular activity. Contraction of the muscles causes intermittent venous compression, and because of the orientation of the venous valves, blood is forced from the veins toward the heart. Therefore, muscular contraction in the lower limb lowers the mean venous pressure and serves as an auxiliary pump to assist venous return. Muscle contraction lowers capillary hydrostatic pressure and increases local blood circulation.

How OnPulse works









The geko and firefly devices are a small disposable, internally powered, neuromuscular stimulation device that is applied externally to the leg. They are self-adhesive and applied to the outer/lateral aspect of the knee. This positioning enables integral electrodes to apply a stimulus to the common peroneal nerve. This nerve controls the contraction of the calf muscles. The stimulation of this nerve by the device causes the muscles to contract isometrically and will not affect normal movement of the limb nor mobility of the patient. Contraction of the calf muscles will boost blood flow from the lower limbs back to the heart, thus increasing venous return, local blood circulation and help prevent venous thrombosis. The device has several stimulation levels (see table below) to balance maximal effect of stimulation with patient comfort. It is fully insulated by the protective moulding and there is no risk of shock

The application of the device is very simple, and the patient will only experience a cooling effect as the area of skin, to which the device will be applied, is cleaned. Thereafter, the patient will feel as if a small adhesive patch has been applied to the skin. Upon switching on the device and selecting the appropriate stimulation level, the patient will be aware of the muscle contraction, awareness of which will recede slightly after a few minutes (accommodation). Over the next hour and the treatment period the patient's awareness of muscle contraction will lessen, and the patient can carry out their normal routine including sleep.


Be sure that the device is removed if the patient needs to shower or bathe.



Device output specifications

| Product name | geko™ | | | | | | firefly™ | |
|-------------------|---|---|---|---|---|---|---|---|
| Model reference | T-1 | T-2 | W-2 | T-3 | XT-3 | W-3 | T-1 | T-2 |
| Model Identifier |  |  |  |  |  |  |  |  |
| Product type | Powered muscle stimulator | | | | | | | |
| Class | BF (The entire device is considered to be a patient applied part) | | | | | | | |
| Dimensions | 149mm x 42mm x 11mm | 186mm x 31mm x 9.35mm | 186mm x 31mm x 9.35mm | 186mm x 31mm x 9.35mm | 186mm x 31mm x 9.35mm | 186mm x 31mm x 9.35mm | 149mm x 42mm x 11mm | 186mm x 31mm x 9.35mm |
| Weight | 16g (device only) | 10g (device only) | 10g (device only) | 10g (device only) | 10g (device only) | 10g (device only) | 16g (device only) | 10g (device only) |
| Power source | Internally powered equipment, battery not replaceable | | | | | | | |
| Battery | Primary lithium coin cell - removable for disposal | | | | | | | |
| Operation | continuous operation – equipment not suitable for use in presence of flammable anaesthetic mixture with air or with oxygen or nitrous oxide | | | | | | | |
| Stimulation modes | 7 (selected pulse widths) | 7 (selected pulse widths) | 10 (selected pulse widths) | 11 (selected pulse) | 11 (selected pulse) | 10 (selected pulse widths) | 7 (selected pulse widths) | 7 (selected pulse widths) |
| Pulse current | 27mA | 27mA | 54mA | 27, 38 or 54mA | 27, 38 or 54mA | 27, 38 or 54mA | 27mA | 27mA |
| Load impedance | 200Ω to 3kΩ for 27mA output | 200Ω to 3k Ω for 27mA output | 200Ω to 3k Ω for 54mA output | 200Ω to 3k Ω for 54mA output | 200Ω to 3k Ω for 54mA output | 200Ω to 3k Ω for 54mA output | 200Ω to 3kΩ for 27mA output | 200Ω to 3k Ω for 27mA output |
| Pulse voltage | Set by current and load | | | | | | | |
| Pulse width ±10% | 70, 100, 140, 200, 280, 400 and 560µs | 50, 70, 100, 140, 200, 280, 400µs | 25, 35, 50, 70, 100, 140, 200, 280, 400 and 560µs | 35, 50, 70, 100, 140, 200, 280 µs @27mA, 280 & 400µs @38mA, 400, 560 µs @54mA, | 35, 50, 70, 100, 140, 200, 280 µs @27mA, 280 & 400µs @38mA, 400, 560 µs @54mA, | 50, 70, 100, 140, 200 µs @27mA, 200, 280 & 400µs @38mA, 400, 560 µs @54mA, | 70, 100, 140, 200, 280, 400 and 560µs | 50, 70, 100, 140, 200, 280, 400µs |
| Repetition rate | 1HZ (±5%) | | | | | | | |
| Maximum charge | 20µC per pulse | 20µC per pulse | 40µC per pulse | 40µC per pulse | 40µC per pulse | 40µC per pulse | 20µC per pulse | 20µC per pulse |
| Net charge output | Less than 0.1 µC per cycle. Charge balance is provided by return pulses of low amplitude and same the same total charge as the stimulation pulse | | | | | | | |
| Output coupling | Ceramic capacitor | | | | | | | |
| Operating time | 24 hour duration (maximum 30 hours) | 24 hour duration (maximum 30 hours) | Two, 6 hour sessions | 24 hour duration (maximum 30 hours) | 24 hour duration (maximum 30 hours) | Two, 12 hour sessions | 24 hour duration (maximum 30 hours) | 24 hour duration (maximum 30 hours) |
| Mode of Operation | The devices are suitable for continuous operation | | | | | | | |
| Indicator display | green LED , flashing to indicate operation and the setting level (pulse width): the number of flashes in the sequence | | | | | | | |
| fault indication | the stimulator device will automatically switch off for over current, under current, low battery voltage or end of operating time | | | | | | | |

geko™, firefly™ and OnPulse™ are trademarks of Sky Medical Technology Ltd.



Output voltages and currents: measured at internal outputs of the pulse generator ($\pm 15\%$)

geko™ T-1 and firefly™ T-1

| pulse width | half-power setting 280 μ s | | full-power setting 560 μ s | |
|-----------------------------|--------------------------------|---------|--------------------------------|---------|
| load | current | voltage | current | voltage |
| 200 Ω | 27mA | 5.4V | 27mA | 5.4V |
| 500 Ω | 27mA | 13.5V | 27mA | 13.5V |
| 1000 Ω | 27mA | 27V | 27mA | 27V |
| 2000 Ω | 27mA | 54V | 27mA | 54V |
| 3000 Ω | 27mA | 81V | 27mA | 81V |
| current rms (500 Ω) | 1mA rms maximum | | | |
| voltage rms (500 Ω) | 0.5V rms maximum | | | |

geko™ T-2 and firefly™ T-2

| pulse width | half-power setting 200 μ s | | full-power setting 400 μ s | |
|-----------------------------|--------------------------------|---------|--------------------------------|---------|
| load | current | voltage | current | voltage |
| 200 Ω | 27mA | 5.4V | 27mA | 5.4V |
| 500 Ω | 27mA | 13.5V | 27mA | 13.5V |
| 1000 Ω | 27mA | 27V | 27mA | 27V |
| 2000 Ω | 27mA | 54V | 27mA | 54V |
| 3000 Ω | 27mA | 81V | 27mA | 81V |
| current rms (500 Ω) | 1mA rms maximum | | | |
| voltage rms (500 Ω) | 0.5V rms maximum | | | |

geko™ W-2

| pulse width | half-power setting 280µs | | full-power setting 560µs | |
|--------------------|--------------------------|---------|--------------------------|---------|
| load | current | voltage | current | voltage |
| 200Ω | 54mA | 10.8 V | 54mA | 10.8 V |
| 500Ω | 54mA | 27 V | 54mA | 27 V |
| 1000Ω | 54mA | 54 V | 54mA | 54 V |
| 2000Ω | 54mA | 108 V | 54mA | 108 V |
| 3000Ω | 54mA | 162 V | 54mA | 162 V |
| current rms (500Ω) | 1.3mA rms maximum | | | |
| voltage rms (500Ω) | 0.7V rms maximum | | | |

geko™ T-3 and geko™ XT-3

| pulse width | half-power setting 400µs @38mA (level 9) | | full-power setting 560µs @54mA (level 11) | |
|--------------------|--|---------|---|---------|
| load | current | voltage | current | voltage |
| 200Ω | 38mA | 7.6 V | 54mA | 10.8 V |
| 500Ω | 38mA | 19 V | 54mA | 27 V |
| 1000Ω | 38mA | 38 V | 54mA | 54 V |
| 2000Ω | 38mA | 76 V | 54mA | 108 V |
| 3000Ω | 38mA | 114 V | 54mA | 162 V |
| current rms (500Ω) | 1.3mA rms maximum | | | |
| voltage rms (500Ω) | 0.7V rms maximum | | | |

geko™ W-3

| pulse width | half-power setting 280µs @54mA (level 8) | | full-power setting 560µs @54mA (level 10) | |
|--------------------|--|---------|---|---------|
| load | current | voltage | current | voltage |
| 200Ω | 54mA | 10.8 V | 54mA | 10.8 V |
| 500Ω | 54mA | 27 V | 54mA | 27 V |
| 1000Ω | 54mA | 54 V | 54mA | 54 V |
| 2000Ω | 54mA | 108 V | 54mA | 108 V |
| 3000Ω | 54mA | 162 V | 54mA | 162 V |
| current rms (500Ω) | 1.3mA rms maximum | | | |
| voltage rms (500Ω) | 0.7V rms maximum | | | |

geko™ & firefly™

| | |
|-------------------|--|
| Indicator display | green LED , flashing fault indication: the stimulator device will automatically switch off for over current, under current, low battery voltage or end of its run time (see the Instructions for Use provided with the device for details of the run (treatment) time). |
| Standards | EN60601-1:2006, IEC 60601-2-10:2012, EN60601-1-11:2010, EN 60601-1-2:2015, ISO 10993-1:2009 |

Operating conditions:

| | |
|-----------------------|-----------------------------|
| Temperature range | 5°C to 40°C, |
| Humidity range | up to 93% RH non-condensing |
| Atmospheric pressures | 70 kPa to 106 kPa |

The device should be used at between 5 and 40°C. If the storage conditions were outside of this range allow time for the device to reach room temperature before use: 30 minutes should be sufficient.

The device is compatible with HF surgical equipment and will continue to operate in its presence. The device must be outside of the sterile area and it is recommended that electrodes are not be used within 50mm from the device in order to prevent inadvertent damage to the geko device.

Storage & transport conditions in original packaging:

Store the device in its protective foil pouch. The storage and temperature range is -25°C to 40°C, though the device will still operate after exposure to 70°C for short durations (up to 2 weeks in total). As it is difficult to monitor storage and transportation temperature, we recommend to store the device at room temperature (up to 30°C), if at all possible.

| | |
|-----------------------|---|
| Temperature range | -25°C to 40°C (up to 70°C for short durations only) |
| Humidity range | up to 93% RH non-condensing |
| Atmospheric pressures | 70 kPa to 106 kPa |
| Shelf-life | see expiry date on the pouch label |

Storage conditions outside of original packaging:

The device will deteriorate if removed from its original package and must be left in its original package until just prior to use. The device may be removed temporarily and reapplied if necessary, for example to prevent the device getting wet during bathing or showering. If it must be removed keep it at clean, dry and at room temperature between uses, and reapply as soon as possible.

Use-life

| | |
|-----------------------|---|
| firefly™ T-1 | Up to 30 hours of operational life |
| geko™ T-1 | Up to 30 hours of operational life in continuous use |
| geko™ T-2 | Up to 30 hours of operational life in continuous use |
| geko™ W-2 | Up to 12 hours of operational life, in two, 6 hour treatment sessions |
| geko™ W-3 | Up to 24 hours of operational life, in two, 12 hour treatment sessions |
| geko™ T-3, geko™ XT-3 | 30 hours of operational life from the time it is first turned on. After 30 hours has elapsed, the device will be permanently disabled |

Materials:

geko™ T-1, firefly™ T-1 Polyethylene terephthalate (Mylar), ethylene-vinyl acetate, polypropylene, hydrogel
geko™ T-2, W-2, T-3, XT-3, W-3, firefly™ T-2 Polyethylene terephthalate (Mylar), polypropylene, polyether ester, hydrogel

Side Effects**Skin Inflammation or Irritation**

In some cases, skin inflammation or irritation can develop in the contact area: either remove the device or re-attach in the alternative fitting positions. If the condition persists or recurs, obtain specialist medical advice before resuming use.

Warnings**Seek specialist medical opinion if the patient is/has:**

- Implanted electronic devices (e.g. a cardiac pacemaker),
- Recently diagnosed or suspected DVT,
- Pregnant or breast feeding,
- Diagnosed heart condition,
- Epilepsy,
- Had recent surgery where muscle contractions may disrupt the healing process,
- Used the device for 28 consecutive days,
- Allergic to acrylic acid (the gel contains acrylic polymers)

Do not use:

- During any activity in which involuntary muscle contractions may put the user or others at risk of injury (e.g. Driving or operating machinery)
- When bathing or showering – switch off the device and remove temporarily
- If the device has been worn by another individual – this will carry risk of infection
- If packaging is open or damaged
- If device is damaged

Device should not be used on the following areas of the body:

- Head
- Eyes
- Mouth
- Neck (especially the carotid sinus)
- On the chest, upper back or crossing over the heart. This may increase the risk of cardiac fibrillation.

Do not apply over or near the following:

- Sore,
- Infected or inflamed areas,
- Broken skin or skin eruptions, e.g. phlebitis, thrombophlebitis, varicose veins etc.
- Cancerous lesions.

Do not use in proximity of the following equipment/environments, which could result in the possible degradation of the performance of the geko device:

- Short wave/ microwave equipment (i.e. within 1m)
- Portable RF communications equipment (including peripherals such as antenna cables and external antennas) (i.e. within 30cm)
- Heat sources, such as fires or radiant heaters

Do not use in oxygen rich environments.

Use on children - the safety of the device has not been tested in children; we do not recommend using the device on children.

Precautions

Keep out of the reach of children and pets.

Do not place the geko device in the mouth; it is a choke and potential allergic hazard. If the device or any component is swallowed seek IMMEDIATE medical assistance.

Serious harm could be caused if the battery is swallowed.

No modification of this equipment is allowed.

Excessive force may damage the device.

MRI -The device should be removed before the patient undergoes MRI as it contains ferromagnetic components.

ECG -The device should be switched off during ECG monitoring using leg electrodes as it may interfere with ECG leg electrode signals.

Any serious incident that has occurred in relation to the device should be reported to the manufacturer and the competent authority of the Member State in which the user and/or patient is established.

Visit www.gekodesices.com for further advice and usage on the geko™ devices.

Visit www.fireflyrecovery.com for further advice and usage on the firefly™ devices.

Identification

Each device has a model identification logo which can be seen on the device itself, on the labels and the Instructions for Use provided with the device. See above for details. The label bears the manufacturing Lot number and expiry date of the device.

Instructions for Use

Please read the Instructions for Use BEFORE using the device. Both this document and the Instructions for Use provided with the device include important fitting instructions and safety information essential for safe and effective use of the device.

IP Classification


















The device has an Ingress Protection (IP) rating of IP22. See the symbols section below for further explanation.

Electro-Magnetic Compatibility (EMC)

The devices have been tested to current safety standards for emission of, and immunity from electro-magnetic radiation, and found to comply. See EMC Declaration below, for full details.

The devices have been tested to current safety standards for immunity from Electrostatic Discharge (ESD). In some circumstances a device may turn off when subject to ESD: however, the device will be undamaged and will operate normally once switched back on.

Symbols

| | |
|--|---|
|  | Type BF applied part |
|  | Product not manufactured with Latex |
|  | Single use only – use only on one patient for a single course of treatment |
|  -25°C - +40°C | Storage and transportation atmospheric pressure range whilst within packaging |
|  | Lot number |
|  | Catalogue number |
|  | Expiry date – do not use after this date |
|  70kPa ↔ 106kPa | Storage and transportation pressure range whilst within packaging |
|  | Ingress protection rating 22 |
|  | EU Authorised Representative |
|  | Device identifier |
|  0% - 93% | Storage and transportation humidity range whilst within packaging |
|  | Manufactured by |
|  | See Instructions for Use |
|  | CE Mark of Conformity |
|  | Do not use if package is damaged |
|  | EU Importer |

Declaration of Conformity – Electromagnetic Compatibility

Guidance and manufacturer's declaration – electromagnetic emissions

The geko and firefly devices are intended for use in the electromagnetic environment specified below. The customer or the user of the device should assure that it is used in such an environment

| Emissions test | Compliance | Electromagnetic environment-guidance |
|---|----------------|--|
| RF emissions CISPR 11 | Group 1 | The devuces uses RF energy only for its internal function. Therefore, its RF emissions are very low and are not likely to cause any interference in nearby electronic equipment. |
| RF emissions CISPR 11 | Class B | The devices are suitable for use in all establishments, including domestic establishments and those directly connected to the public low-voltage power supply network that supplies building used for domestic purposes. The devices are internally powered by a lithium coin cell CR2032 |
| Harmonic emissions IEC 61000-3-2 | Not applicable | |
| Voltage fluctuations/ flicker emissions IEC 61000-3-3 | Not applicable | |

Guidance and manufacturer's declaration – electromagnetic immunity

The geko and firefly are intended for use in the electromagnetic environment specified below. The customer or the user of the device should assure that it is used in such an environment

| Immunity test | IEC60601-1-2 test level | Compliance level | Electromagnetic environment-guidance |
|--|---|---------------------------------------|---|
| Electrostatic discharge IEC 61000-4-2 | ±8 kV contact ±2, 4, 8, 15 kV air | Not applicable ±2, 4, 8, 15 kV air | Floors should be wood, concrete or ceramic tile. If floors are covered with synthetic material, the relative humidity should be at least 30%. The device may turn-off following Electrostatic discharge, however the device will be undamaged and will operate normally after re-starting |
| Electrical fast transient/burst IEC 61000-4-2 | ±2 kV for power supply lines ±1 kV for input / output lines | Not applicable | Not applicable as internally powered device |
| Surge IEC61000-4-5 | ±1 kV differential mode ±2 kV common mode | Not applicable | Not applicable as internally powered device |
| Voltage dips, short interruptions and voltage variations on power supply lines IEC6100-4-11 | <5% U_T (>95% dip in U_T) for 0.5 cycle 40% U_T (60% dip in U_T) for 5 cycles 70% U_T (30% dip in U_T) for 25 cycles <5% U_T (>95% dip in U_T) for 5 seconds | Not applicable | Not applicable as internally powered device |
| Conducted immunity | | Not applicable | No mains cable or external connections |
| Power frequency (50/60Hz) Magnetic field IEC61000-4-8 | 30 A/m | 30 A/m | Power frequency magnetic fields should be at levels characteristic of a typical location in a typical commercial or hospital environment |

Note U_T is the a.c. mains voltage prior to the application of the test level

Guidance and manufacturer's declaration – electromagnetic immunity

The geko and firefly are intended for use in the electromagnetic environment specified below. The customer or the user of the device should assure that it is used in such an environment

| Immunity test | IEC60601-1-2 test level | Compliance level | Electromagnetic environment-guidance |
|----------------------------|--------------------------|------------------|---|
| | | | Portable and mobile RF communications equipment should be no closer to any part of the devices, including cables, than the recommended distances calculated from the equation applicable to the frequency of the transmitter Recommended separation distance |
| Conductive RF IEC61000-4-6 | 3 V rms: 150kHz to 80MHz | 3 V rms | $d=1.2\sqrt{P}$ |
| Radiated RF IEC61000-4-3 | 10 V/m: 80MHz to 2,5GHz | 10 V/m | $d=1.2\sqrt{P}$ 80MHz to 800MHz $d=2.3\sqrt{P}$ 800MHz to 2.5GHz |
| | | | Where P is the maximum output power rating of the transmitter in watts (W) according to the transmitter manufacturer and d is the recommended separation distance in metres(m). Field strengths from fixed RF transmitters, as determined by an electromagnetic survey, ^a should be less than the compliance level in each frequency range. ^b Interference may occur in the vicinity of equipment marked with the following symbol: |



Note 1 At 80MHz and 800MHz the higher frequency range applies.

Note 2 These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

^a Field strengths from fixed transmitted, such as base stations for radio (cellular/cordless) telephones and land mobile radios, amateur radio, AM and FM radio broadcast cannot be predicted theoretically with accuracy. To assess the electromagnetic environment in the location due to fixed RF transmitters, an electromagnetic site survey should be considered. If the measured field strength in the location in which the device is used exceeds the applicable RF compliance level above, the devices should be observed to verify normal operation. If abnormal performance is observed, additional measures may be necessary, such as reorienting or relocating the devices.

^b Over the frequency range 150kHz to 80MHz, field strengths should be less than 3V/m.

Recommended separation distances between portable and mobile communication equipment and the geko and firefly devices

The geko and firefly are intended for use in an electromagnetic environment in which radiated RF disturbances are controlled. The customer or the user of the device can help prevent electromagnetic interference by maintaining a minimum distance between portable and mobile RF communication equipment (transmitters) and the devices as recommended below, according to the maximum output power of the communications equipment.

| Rated maximum output power of transmitter W | Separation distance according to frequency of transmitter (m) | | |
|---|---|------------------------------------|-------------------------------------|
| | 150kHz to 80MHz $d=1.2\sqrt{P}$ | 80MHz to 800MHz $d=1.2\sqrt{P}$ | 800MHz to 2.5GHz $d=2.3\sqrt{P}$ |
| 0.01 | 0.12 | 0.12 | 0.23 |
| 0.1 | 0.38 | 0.38 | 0.73 |
| 1.0 | 1.2 | 1.2 | 2.3 |
| 10 | 3.8 | 3.8 | 7.3 |
| 100 | 12 | 12 | 23 |

For transmitter rated at a maximum output power not listed above, the recommended separation distance d in meters (m) can be estimated using the equation applicable to the frequency of the transmitter, where P is the maximum output power of the transmitter in watts (W) according to the transmitter manufacturer.

Note 1. At 80MHz and 800MHz the higher frequency range applies.

Note 2. These guidelines may not apply in all situations. Electromagnetic propagation is affected by absorption and reflection from structures, objects and people.

Test specification for ENCLOSURE PORT IMMUNITY to RF wireless communications equipment

| Test frequency (MHz) | Band ^{a)} (MHz) | Service ^{a)} | Modulation ^{b)} | Maximum power (W) | Distance (m) | IMMUNITY TEST LEVEL (V/m) |
|----------------------|--------------------------|--|---|-------------------|--------------|---------------------------|
| 385 | 380 - 390 | TETRA 400 | Pulse modulation ^{b)} 18Hz | 1,8 | 0,3 | 27 |
| 450 | 430 - 470 | SMRS 460, FRS 460 | FM ^{c)} ± 5 kHz deviation 1 kHz sine | 2 | 0,3 | 28 |
| 710 | 704 - 787 | LTE Band 13, 17 | Pulse modulation ^{b)} 217 Hz | 0,2 | 0,3 | 9 |
| 745 | | | | | | |
| 780 | | | | | | |
| 810 | 800 - 960 | GSM 800/900, TETRA 800, iDEN 820, CDMA 850, LTE Band 5 | Pulse modulation ^{b)} 18 Hz | 2 | 0,3 | 28 |
| 870 | | | | | | |
| 930 | | | | | | |
| 1 720 | 1 700 – 1 990 | GSM 1800; CDMA 1900; GSM 1900; DECT; LTE Band 1, 3, 4, 25; UMTS | Pulse modulation ^{b)} 217 Hz | 2 | 0,3 | 28 |
| 1 845 | | | | | | |
| 1 970 | | | | | | |
| 2 450 | 1 400 – 2 570 | Bluetooth, WLAN, 802.11 b/g/n, RFID 2450, LTE Band 7 | Pulse modulation ^{b)} 217 Hz | 2 | 0,3 | 28 |
| 5 240 | 5 100 – 5 800 | WLAN 802.11 a/n | Pulse modulation ^{b)} 217 Hz | 0,2 | 0,3 | 9 |
| 5 500 | | | | | | |
| 5 785 | | | | | | |

NOTE If necessary to achieve the IMMUNITY TEST LEVEL, the distance between the transmitting antenna and the ME EQUIPMENT or ME SYSTEM may be reduced to 1 m. The 1 m test distance is permitted by IEC 61000-4-3.

a) For some services, only the uplink frequencies are included.

b) The carrier shall be modulated using a 50 % duty cycle square wave signal.

c) As an alternative to FM modulation, 50 % pulse modulation at 18 Hz may be used because while it does not represent actual modulation, it would be worst case.