



**USULTRA, LLC**  
Certificate of Analysis / QC Results

<b>Packaged Product:</b>	ASTM2-55G DEIONIZED WATER ASTM TYPE II - 55 GALLON		
<b>Test</b>	<b>Target/UOM</b>	<b>Specification</b>	<b>Result</b>
APPEARANCE (CLEAR, NO INSOLUBLES)	PASS	Y/N	PASS
AEROBIC BACTERIA	PASS	<10 CFU/100ML	PASS
ENDOTOXIN (KINETIC)	PASS	<0.25 EU/ML	PASS
SILICATES	0.00 PPB	<3.00 PPB	PASS
TOTAL ORGANIC CARBON	0 PPB	<50 PPB	PASS
RESISTIVITY (22°C - 30°C)	>10 megohm	>10 megohm	PASS
CONDUCTIVITY (22°C - 30°C)	0.1 MS/CM	0.1 MS/CM	PASS
Lot # 0001312 Made: 23 Aug 2023 Expires: 23 Feb 2025			