



USULTRA, LLC
Certificate of Analysis / QC Results
Lot: 04032023326-500

Packaged Product:		CUO2122-00500 VERIF.1 PURIFIED WATER USP--PW-500ML--K6	
Test	Target/UOM	Range	Result
APPEARANCE (CLEAR, NO INSOLUBLES)	Y PASS	Y/N	Y
ENDOTOXIN (KINETIC)(<0.25EU/ML)	Y PASS	Y/N	Y
PH (20°C)	7.5 PH	6.5 - 8.5	7.00
TOTAL ORGANIC CARBON	0 PPB	0-10	4
CONDUCTIVITY (22°C - 30°C)	0 MS/CM	0.00-1.30	0.05
Lot: 04032023326-500 Made: 4/3/23 Expires: 10/2/24			