



USULTRA, LLC
Certificate of Analysis / QC Results

Packaged Product:	ASTM1-20L DEIONIZED WATER ASTM TYPE I - 20 LITER		
Test	Target/UOM	Range	Result
APPEARANCE (CLEAR, NO INSOLUBLES)	PASS	Y/N	PASS
AEROBIC BACTERIA	PASS	<1CFU/1ML	PASS
ENDOTOXIN (KINETIC)	PASS	<0.25 EU/1ML	PASS
PARTICULATE	PASS	<0.2 μ	PASS
TOTAL ORGANIC CARBON	0 PPB	<10 PPB	PASS
RESISTIVITY (22°C - 30°C)	> 18 megohm	>18 megohm	PASS
CONDUCTIVITY (22°C - 30°C)	<0.056 MS/CM	<0.056 MS/CM	PASS
Lot #: 06092023215-20 Made: 06/09/2023 Expires: 12/09/2024			