



**USULTRA, LLC**  
Certificate of Analysis / QC Results

<b>Packaged Product:</b>	ASTM1-20L DEIONIZED WATER ASTM TYPE I - 20 LITER		
<b>Test</b>	<b>Target/UOM</b>	<b>Range</b>	<b>Result</b>
APPEARANCE (CLEAR, NO INSOLUBLES)	PASS	Y/N	PASS
AEROBIC BACTERIA	PASS	<1CFU/1ML	PASS
ENDOTOXIN (KINETIC)	PASS	<0.25 EU/1ML	PASS
PARTICULATE	PASS	<0.2 $\mu$	PASS
TOTAL ORGANIC CARBON	0 PPB	<10 PPB	PASS
RESISTIVITY (22°C - 30°C)	> 18 megohm	>18 megohm	PASS
CONDUCTIVITY (22°C - 30°C)	<0.056 MS/CM	<0.056 MS/CM	PASS
Lot #: 20240328 Made: 2024-03-28 Expires: 2025-09-28			