



SIRIM Berhad
Industrial Biotechnology Research Centre, Building 19
Tel: 03-55446953/6960
Fax: 03-55446988

TEST REPORT

EVALUATION OF EYE IRRITATION ON QUANTUM ION USING *IN VITRO* RECONSTRUCTED HUMAN CORNEA-LIKE EPITHELIUM TISSUE MODEL EPIOULAR™ EYE IRRITATION TEST

Job No. J736/20

Report No. R736/20/B19/44

Sponsor:

Eva Energy Sdn Bhd,
12, Jalan Bandar 20,
Pusat Bandar Puchong,
47160 Puchong,
Selangor

Test Facility:

Industrial Biotechnology Research Centre (IBRC),
Building 19, SIRIM Berhad

Study Initiation Date:

29 June 2020

Experimental Start Date:

05 August 2020

Experimental End Date:

11 August 2020

Study Completion Date:

18 August 2020

SIRIM Berhad

(No. Syarikat 367474 - V)
1, Persiaran Dato' Menteri
Seksyen 2, Peti Surat 7035
40700 Shah Alam
MALAYSIA
Tel : 60-3-55446000
Hotline : 60-3-55103535
Fax : 60-3-55108095
Website : www.sirim.my

TEST REPORT

REPORT NO: R736/20/B19/44

PAGE: 2 of 18

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

APPROVED SIGNATORIES

We, the undersigned, declare that the methods, results and data contained in this report faithfully reflect the procedures used and raw data collected throughout the study.



(SUZAINI BADRUDIN)

Reviewer

Industrial Biotechnology Research Centre

26 AUG 2020

Date



(NURHAYATI ARIFFIN)

Analyst

Industrial Biotechnology Research Centre

26 AUG 2020

Date



MS ISO/IEC 17025
TESTING
SAMM NO.676

TEST REPORT

REPORT NO: R736/20/B19/44

PAGE: 3 of 18

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

SUMMARY

EVALUATION OF EYE IRRITATION ON QUANTUM ION USING *IN VITRO* EPIOCULAR™ EYE IRRITATION TEST (EIT)

In vitro eye irritation test on Quantum Ion using *in vitro* EpiOcular™ Eye Irritation Test (EIT) was performed according to Standard Operating Procedure developed at MatTek Corporation. This *in vitro* standard method was validated by European Centre of the Validation of Alternative Methods (ECVAM) as *in vitro* test method based on reconstructed human cornea-like epithelium (RhCE) technology. The test was conducted in line with the requirement of OECD Guidelines for Testing of Chemicals No 492.

The test was conducted to determine whether the test item causes irritation to the *in vitro* eye model EpiOcular™.

In vitro EpiOcular™ eye irritation test assessed irritancy level of the Quantum Ion via topical exposure of the sample on three-dimensional reconstructed human cornea-like epithelial (RhCE) model EpiOcular™, followed by cell viability test. The test was conducted in duplicate; two RhCE tissues. After (30 ± 2) minutes of exposure, tissues were thoroughly rinsed, blotted to remove the test item and followed by (12 ± 2) minutes post-treatment immersion (Post-Soak) at room temperature. Then, the medium was changed and tissues were incubated for another (120 ± 15) minutes. Next, MTT assay was performed by transferring the tissues into 24-well plates containing MTT (3-[4,5-dimethylthiazol-2-yl]-2,5-diphenyltetrazolium bromide) medium (1 mg/mL). After (180 ± 10) minutes of incubation, blue formazan salt formed via chemical conversion of MTT in cellular mitochondria was extracted with 2.0 mL isopropanol / tissue. Optical density of the extracted formazan was determined at 570 nm using spectrophotometer. Relative cell viability was calculated for each tissue as % of the mean of the negative control tissues. The eye irritation potential was classified according to the remaining cell viability obtained after test item treatment.

The data indicates that Quantum Ion did not reduce viability of the EpiOcular™ tissue to below 60 % of the negative control. Under the condition of this test, Quantum Ion is considered as **Non-Irritant** to *in vitro* eye model EpiOcular™.



MS ISO/IEC 17025
TESTING
SAMM NO.676

TEST REPORT

REPORT NO: R736/20/B19/44

PAGE: 4 of 18

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

BACKGROUND

According to United Nations Globally Harmonized System of Classification and Labelling of Chemicals (UN GHS), eye irritation refers to the production of changes in the eye following application of a test chemical to the anterior surface of the eye, which are fully reversible within 21 days of application. The EpiOcular Eye Irritation Test predicts the acute eye irritation potential of a topically applied chemical or formulation by measurement of its cytotoxic effect of the EpiOcular cornea epithelial model.

The EpiOcular™ OCL-200 Eye Irritation Test (OCL-200-EIT) RhCE tissue construct is similar to the *in vivo* corneal epithelium three-dimensional structure and is produced using cells from the species of interest. The EpiOcular™ tissue construct is a non-keratinized epithelium prepared from normal human keratinocytes. It offers features appropriate for a model of ocular irritation.

Based on the depth of injury model, the EpiOcular™ Eye Irritation Test is intended to differentiate those materials that are non-irritants (would not require a warning label in the European chemical classification systems) from those that would require labeling as either Globally Harmonized System (GHS) 1 or 2. Liquids and solids are treated with different exposure and post-exposure incubations. Minimum of two construct tissues are used for each treatment and control group. Relative tissue viability is determined against the negative control-treated constructs by the NAD(P)H-dependent microsomal enzyme reduction of MTT (3-[4,5 - dimethylthiazol-2-yl] - 2,5 - diphenyltetrazolium bromide (and to a lesser extent, by the succinate dehydrogenase reduction of MTT) in control and test article-treated cultures (Berridge, et al., 1996). Thus, the toxicity of the test item or the ocular irritation potential is evaluated by the relative viability of the treated tissues relative to the negative control-treated tissues.

From validation study and its independent peer review, it was concluded that the EpiOcular™ EIT able to correctly identify chemicals (both substances and mixtures), not requiring classification and labelling for eye irritation or serious eye damage according to UN GHS, and the test method was recommended as scientifically valid for that purpose. The EpiOcular™ EIT is thus referred to as the Validated Reference Method (VRM) in the present Test Guideline.



MS ISO/IEC 17025
TESTING
SAMM NO.676

TEST REPORT

REPORT NO: R736/20/B19/44

PAGE: 5 of 18

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

CONTENTS

APPROVAL SIGNATURES	2
SUMMARY	3
BACKGROUND	4
1.0 OBJECTIVE	6
2.0 STUDY TIMETABLE	6
3.0 MATERIALS	6
4.0 METHOD	9
5.0 DATA ANALYSIS	10
6.0 ACCEPTABILITY RANGES FOR QUALITY CONTROL	12
7.0 ACCEPTABILITY RANGES FOR NEGATIVE CONTROL	12
8.0 ACCEPTANCE CRITERIA FOR POSITIVE CONTROL	12
9.0 DATA INTERPRETATION PROCEDURE (PREDICTION MODEL)	12
10.0 RESULT AND DISCUSSION	13
11.0 CONCLUSION	15
12.0 RETENTION OF RECORDS AND TEST ITEM	15
13.0 REFERENCES	15
TABLES	
Table 1 Optical Densities (ODs)	16
Table 2 Blank Corrected Data	16
Table 3 Classification of Test Item	17
FIGURES	
Figure 1 Relative Viability	18



MS ISO/IEC 17025
TESTING
SAMM NO.676

TEST REPORT

REPORT NO: R736/20/B19/44

PAGE: 6 of 18

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

1.0 OBJECTIVE

The objective of this study was to predict eye irritation potential of Quantum Ion using an *in vitro* eye irritation test using EpiOcular™ model.

2.0 STUDY TIMETABLE

2.1 Receipt of Reconstructed Human Cornea-like Epithelium (RhCE) model EpiOcular™ tissues:

05 August 2020

2.2 Tissue Conditioning

05 August 2020

2.3 Pre-Treatment

06 August 2020

2.4 Treatment

06 August 2020

2.5 Rinsing

06 August 2020

2.6 Post-Soak

06 August 2020

2.7 Post Incubation

06 August 2020

2.8 MTT Viability Test

06 August 2020

2.9 Optical Density Reading

07 August 2020

2.10 Data Analysis

08 August 2020 – 11 August 2020

3.0 MATERIALS

3.1 Test Item

3.1.1 Test item: Quantum Ion

3.1.2 Sample marking: Copper Ion

3.1.3 Date received: 29 June 2020

3.1.4 Physical appearance: Liquid



MS ISO/IEC 17025
TESTING
SAMM NO.676

TEST REPORT

REPORT NO: R736/20/B19/44

PAGE: 7 of 18

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

- 3.1.5 Colour: Colourless
- 3.1.6 Physical Chemical Properties Data: Not provided
- 3.1.7 Quantity received : 250 ml + 50 ml + 50 ml
- 3.1.8 pH: Not provided
- 3.1.9 Storage condition: Room temperature
- 3.1.10 Solubility: Not provided
- 3.1.11 Stability: Not provided
- 3.1.12 Expiration date: Not provided

3.2 Test System

- 3.2.1 Test System: Reconstructed Human Cornea-like Epithelium (RhCE) model EpiOcular™

The Reconstructed Human Cornea-like Epithelium (RhCE) model EpiOcular™ (OCL-200, MatTek, Ashland, USA) consists of normal, human-derived epidermal keratinocytes which have been cultured to form a stratified, squamous epithelium similar to that found in the cornea.

Cultured on specially prepared cell culture inserts using serum-free culture medium, the cells differentiate to form a multi-layered structure which closely parallels the corneal epithelium

- 3.2.2 Product Number: OCL-200 version 2.0
- 3.2.3 Lot No: 28060
- 3.2.4 Production date: 30 July 2020
- 3.2.5 Date of Shipping: 31 July 2020
- 3.2.6 Receipt of EpiOcular™: 05 August 2020
- 3.2.7 Visual quality control of the Eye: All tissues in good condition
- 3.2.8 The EpiOcular™ System is manufactured according to defined quality assurance procedures. All biological components of the EpiOcular™ and the culture medium are tested by manufacturer for viral, bacterial, fungal and mycoplasma contamination. MatTek determines the ET-50 value following exposure to Triton X-100 (0.3%) for each EpiOcular™ lot. The ET-50 must fall within a range established based on a historical database of results or acceptability ranges for quality control based on OECD Guidelines.



MS ISO/IEC 17025
TESTING
SAMM NO.676

TEST REPORT

REPORT NO: R736/20/B19/44	PAGE: 8 of 18
This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.	

3.3 Reagent

3.3.1 Assay Medium: EPI-100-NMM-SIT / Assay Medium

Lot No. : 072720ALB
Sterility : Sterile
Expiration Date : 24/08/2020
Storage : Refrigerator (5 ± 3 °C)
Manufacturer : Mattek Corporation

3.3.2 Phosphate Buffered Saline without Calcium and Magnesium

Lot No.: 070720RAHA
Expiration Date : 07/07/2021
Manufacturer : Mattek Corporation

3.3.3 Phosphate Buffered Saline without Calcium and Magnesium, contain the following:

- i. Sodium Chloride 8 g/L
Lot No.: K50722004912
Manufacturer : MERCK
- ii. Potassium chloride 0.2 g/L
Lot No.: 1409B14J35501
Manufacturer : Bio Basic Canada Inc
- iii. Anhydrous potassium dihydrogen orthophosphate 0.2 g/L
Lot No.: 1211ACN1K12062801
Manufacturer : Bio Basic Canada Inc
- iv. Anhydrous disodium hydrogen orthophosphate 1.15 g/L
Lot No.: 1306ACK2NA12020101
Manufacturer : Bio Basic Canada Inc

3.3.4 MTT – 2mL (5mg/mL)

Lot No. : MKBW0025V
CAS No. : 298-93-1
Manufacturer : SIGMA

3.3.5 MTT Diluent – 8mL

As above 3.3.1

3.3.6 Extractant Solution – Isopropanol

Lot No. : 632261
Storage : Room Temperature
CAS No. : 67-63-0
Manufacturer : Fisher Scientific

3.3.7 Positive Control: Methyl Acetate

Part No. : TC-MA
Lot No. : 070620ALA
Expiration Date : 02/10/2020
Storage : Room Temperature
Manufacturer : MatTek Corporation

3.3.8 Negative Control: Sterile pure water



TEST REPORT

REPORT NO: R736/20/B19/44	PAGE: 9 of 18
This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.	

4.0 METHOD

Upon receipt of the reconstructed human cornea-like epithelial model EpiOcular™, the tissue kits and the solutions were stored according to the manufacturer's directions for unpacking and storage.

4.1 Assessment of Direct Test Item Reduction by MTT

50 µL of the test item was added to 1 mL of the 1 mg/mL MTT and incubated at $(37 \pm 1)^\circ\text{C}$, $(5 \pm 1)\%$ CO₂, 95 % Relative Humidity (RH) for 3 hours. Untreated MTT medium was used as control.

4.2 Assessment of Non-Coloured Material

4.2.1 Isopropanol

50 µL of the test item was added to 2 mL of isopropanol and agitated for 3 hours at room temperature. Isopropanol was run in parallel as negative control.

4.2.2 Sterile pure water

50 µL of the test item was added to 1 mL of water and incubated at $(37 \pm 1)^\circ\text{C}$, $(5 \pm 1)\%$ CO₂, 95 % RH for 60 minutes. Sterile pure water was run in parallel as negative control.

4.3 Tissue Conditioning – Day 0

Under sterile conditions, the sealed 24-well plates containing the EpiOcular™ tissues, was opened. Visual inspection of each insert containing the EpiOcular™ tissue was done prior to tissue conditioning.

1.0 mL of assay medium was dispensed into each well of six-well plates, the EpiOcular™ tissue cultures were transferred into the wells and the plates were incubated for one hour at $(37 \pm 1)^\circ\text{C}$, $(5 \pm 1)\%$ CO₂, 95 % RH. At the end of the first hour of pre-incubation, the assay medium was renewed for further pre-incubation. Each insert was aseptically transferred into well containing 1.0 mL assay medium and pre-incubation was conducted at $(37 \pm 1)^\circ\text{C}$, $(5 \pm 1)\%$ CO₂, 95 % RH for overnight (16 – 24 hours).

4.4 Dosing protocol – Day 1

4.4.1 Pre-Treatment

After overnight incubation, each tissues were pre-wetted with 20 µL of Phosphate Buffered Saline without Calcium and Magnesium (PBS without Ca²⁺ and Mg²⁺). The tissues were incubated at $(37 \pm 1)^\circ\text{C}$, $(5 \pm 1)\%$ CO₂, 95 % RH for 30 ± 2 minutes.

4.4.2 Treatment

50 µl of test item was applied directly on the tissue. Negative and positive controls were run in parallel in an identical way to the dosed cultures. The cultures were incubated at $(37 \pm 1)^\circ\text{C}$, $(5 \pm 1)\%$ CO₂, 95 % RH for 30 ± 2 minutes.



TEST REPORT

REPORT NO: R736/20/B19/44

PAGE: 10 of 18

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

4.4.3 Rinsing

At the end of the exposure period, the EpiOcular™ tissue cultures were removed from the incubator and extensively rinsed with phosphate buffered saline to eliminate any residual test item.

4.4.4 Post-Soak

After rinsing, each EpiOcular™ tissues was immediately immersed in 5 ml of Assay Medium which was previously warmed to room temperature in a pre-labeled 12-well plate. The cultures were incubated at $(37 \pm 1) ^\circ\text{C}$, $(5 \pm 1) \% \text{CO}_2$, 95 % RH for (12 ± 2) minutes.

4.4.5 Post-Incubation

After (12 ± 2) minutes incubation, the medium was changed and EpiOcular™ tissues were incubated in 1ml Assay medium for a further (120 ± 15) minutes at $(37 \pm 1) ^\circ\text{C}$, $(5 \pm 1) \% \text{CO}_2$, 95 % RH.

4.5 MTT Viability Test

4.5.1 The MTT assay was performed by transferring each EpiOcular™ tissues to 24-well plates containing 300 μL MTT medium (1 mg/mL) in each well. After (180 ± 10) minutes of MTT incubation, the blue formazan salt formed by cellular mitochondria was extracted with 2.0 mL isopropanol / tissue. The extraction plates were sealed with parafilm and agitated for at least 2 hours at room temperature.

4.5.2 At the end of the extraction period, EpiOcular™ tissue was pierced with an injection needle and the extract allowed to run into the well from which insert was taken. The insert was then discarded. The extraction solution was pipetted up and down to ensure complete mixing. Finally, 200 μL were transferred into a 96 well microtiter plate for absorbance measurement (OD=optical density) at 570 nm without using a reference filter. 200 μL of isopropanol was used as blank.

Relative cell viability was calculated for each tissue as % of the mean of the negative control tissues.

5.0 DATA ANALYSIS

5.1 Assessment of Direct Test Item Reduction by MTT

The test item is presumed to have reduced the MTT if the MTT solution colour turns blue / purple.

5.2 Assessment of Non- Coloured material

5.2.1 Isopropanol

The test item has to be considered as possibly interacting with the MTT measurement if the OD of the test item solution is more than 0.08 (approximately 5 % of the mean viability of the negative control) after subtraction with the OD of isopropanol. An additional test on colourant controls has to be performed.



TEST REPORT

REPORT NO: R736/20/B19/44

PAGE: 11 of 18

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

5.2.2 Sterile Pure Water

The test item has to be considered as possibly interacting with the MTT measurement if the test item becomes coloured after contact with sterile pure water.

5.3 Workbook EpiOcular-EIT

A blank, password protected MS EXCEL workbook EpiOcular-EIT-SPREAD.XLS was provided by MatTek Corporation. The workbook consists of two spreadsheets named: Import and Spread.

5.4 Raw Data of Optical Densities (ODs)

Raw data of optical densities (ODs) generated by the microplate reader (without blank subtraction) were copied from the reader software and then pasted into the Import spreadsheet of the Excel workbook. The blank corrections, calculation of results and statistical parameters are done automatically in the Spread spreadsheet of the workbook.

5.5 Calculation

5.5.1 After data entry, the spreadsheet performs the following calculations:

5.5.2 Blank correction

5.5.3 For each individual tissue treated with a test substance (TS), the positive control (PC) and the negative control (NC) the individual relative tissue viability was calculated according to the following formulas

$$\text{Relative Viability TS (\%)} = [\text{Corrected OD}_{\text{TS}} / \text{Corrected Mean of OD}_{\text{NC}}] \times 100$$

$$\text{Relative Viability NC (\%)} = [\text{Corrected OD}_{\text{NC}} / \text{Corrected mean of OD}_{\text{NC}}] \times 100$$

$$\text{Relative Viability PC (\%)} = [\text{Corrected OD}_{\text{PC}} / \text{Corrected mean of OD}_{\text{NC}}] \times 100$$

5.5.4 For each test item, negative control and positive control, the mean relative viability of the two individual tissues was calculated and used for classification according to the Prediction Model (Refer to 9.0).



TEST REPORT

REPORT NO: R736/20/B19/44	PAGE: 12 of 18
<p>This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.</p>	

6.0 ACCEPTABILITY RANGES FOR QUALITY CONTROL

	Lower Acceptance Limit	Upper Acceptance Limit
EpiOcular™ EIT (OCL-200)	ET ₅₀ = 12.2 min	ET ₅₀ = 37.5 min

7.0 ACCEPTABILITY RANGES FOR NEGATIVE CONTROL OPTICAL DENSITY (OD) VALUES OF THE TEST METHODS

	Lower Acceptance Limit	Upper Acceptance Limit
EpiOcular™ EIT (OCL-200)	≥ 0.8*	≤ 2.5

*This acceptance limit considers the possibility of extended shipping / storage time (e.g., > 4days) which has been shown not to impact on the performance of the test method.

8.0 ACCEPTANCE CRITERIA FOR POSITIVE CONTROL

The assay meets the acceptance criterion if the mean viability of Positive Control tissues expressed as percent of the negative control tissues is ≤ 50 %. The standard deviation shall be below 20 % for all substances and controls.

	Mean of Viability (%)	Standard Deviation of Viability (SD %)
EpiOcular™ EIT (OCL-200)	≤ 50	<20

9.0 DATA INTERPRETATION PROCEDURE (PREDICTION MODEL)

An irritant is predicted if the mean relative tissue viability of two individual tissues exposed to the test substance is reduced below 60 % of the mean viability of the negative controls.

<i>In vitro</i> result	<i>In vivo</i> prediction
mean tissue viability ≤ 60 %	Irritant (I) (GHS Categories 1 or 2) Further testing in other test method is required
mean tissue viability > 60 %	Non-Irritant (NI) (GHS No Category)



TEST REPORT

REPORT NO: R736/20/B19/44

PAGE: 13 of 18

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.

10.0 RESULT AND DISCUSSION

10.1 Assessment of Direct Test Item Reduction by MTT

There was no change in MTT colour therefore the test item did not interact with MTT.

10.2 Assessment of Non- Coloured material

10.2.1 Isopropanol

There was no colour change and the OD of the test item solution in isopropanol less than 0.08 (approximately 5 % of the negative control). No further testing on colourant controls has to be performed.

10.2.2 Sterile Deionized Water

There was no colour change therefore no additional test has to be performed.

10.3 Viability Measurement

10.3.1 Raw data of optical densities (ODs) of Blank

Refer to Table 1

10.3.2 Blank Correction

Refer to Table 2

10.4 The Quality Control value meets the acceptance range criteria

EpiOcular™ EIT (OCL-200)	Minute
ET ₅₀	17.18

10.5 The negative control OD value meets the acceptance range criteria

EpiOcular™ EIT (OCL-200)	Lower OD	Upper OD
Negative Control	1.358	1.483

10.6 The positive control OD value meets the acceptance range criteria

EpiOcular™ EIT (OCL-200)	Mean of Viability (%)	Standard Deviation of Viability (SD %)
Positive Control	39.1	3.67



MS ISO/IEC 17025
TESTING
SAMM NO.676

TEST REPORT

REPORT NO: R736/20/B19/44	PAGE: 14 of 18
<p>This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.</p>	

10.7 Classification

Test substance, negative control and positive control are qualified according to prediction model in clause 9.0. The mean relative viability of the two individual tissues was calculated and used for classification.

Refer to Table 3

10.8 Graph

10.8.1 The spreadsheet shows a graph of the results (% of relative viability \pm standard deviation)

Refer to Figure 1

10.9 Prediction Model

	Mean of Viability (%)	Standard Deviation of Viability (SD %)	In vitro result	In vivo prediction
Quantum Ion	90.5	3.81	mean tissue viability > 60 %	Non-Irritant (NI)
Negative Control	100.0	6.22	mean tissue viability > 60 %	Non-Irritant (NI)
Positive Control	39.1	3.67	mean tissue viability \leq 50 %	Irritant (I)



TEST REPORT

REPORT NO: R736/20/B19/44	PAGE: 15 of 18
This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.	

11.0 CONCLUSION

Under the condition of this test, Quantum Ion is considered as **Non-Irritant** to *in vitro* eye model EpiOcular™.

12.0 RETENTION OF RECORDS AND TEST ITEM

One report will be forwarded to the Sponsor. The other report, together with all generated raw data is maintained at the Industrial Biotechnology Research Centre Archives.

13.0 REFERENCES

- 13.1 OECD (2015), Reconstructed human Cornea-like Epithelium (RhCE) test method for identifying chemicals not requiring classification and labelling for eye irritation or serious eye damage –OECD Guidelines for Testing of Chemicals No. 492
- 13.2 Protocol for: In *Vitro* EpiOcular™ Eye Irritation Test (OCL-200-EIT) – Reconstructed Human Cornea-like Epithelium (RhCE) Model EpiOcular (OCL-200-EIT). For use with MatTek Corporation
- 13.3 (LWI-238-44) - In Vitro EpiOcular™ Eye Irritation Test (OCL-200-EIT)



TEST REPORT

REPORT NO: R736/20/B19/44	PAGE: 16 of 18
This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.	

Table 1 Optical Densities (ODs) of Blank

Optical Densities (ODs) of Blank	Mean Optical Densities (ODs) of Blank
0.0441	0.0406
0.0427	
0.0408	
0.0397	
0.0394	
0.0390	
0.0399	
0.0391	

Table 2 Blank Corrected Data

	Tissue	Raw data		Blank corrected data		Mean	% of Viability
		1	2	1	2		
Quantum Ion	1	1.3296	1.3364	1.289	1.296	1.292	91.0
	2	1.3204	1.3188	1.280	1.278	1.279	90.0
Negative Control	1	1.5246	1.5224	1.484	1.482	1.483	104.4
	2	1.3921	1.4049	1.352	1.364	1.358	95.6
Positive Control	1	0.5524	0.5652	0.512	0.525	0.518	36.5
	2	0.6379	0.6272	0.597	0.587	0.592	41.7



TEST REPORT

REPORT NO: R736/20/B19/44	PAGE: 17 of 18
<p>This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.</p>	

Table 3 Classification of Test Item

	Mean of OD	SD of OD	Mean of Viability [%]	SD of Viability	CV [%]	In vitro result	Classification
Quantum Ion	1.286	0.013	90.5	3.81	4.21	Non-Irritant	Qualified
Negative Control	1.420	0.125	100.0	6.22	6.22	Non-Irritant	Qualified
Positive Control	0.555	0.074	39.1	3.67	9.39	Irritant	Qualified

OD- Optical Density, SD- Standard Deviation, CV- Coefficient of Variation



MS ISO/IEC 17025
TESTING
SAMM NO. 676

TEST REPORT

REPORT NO: R736/20/B19/44

PAGE: 18 of 18

This report is NOT a Quality Assurance Certificate NOR an Approval Permit. This report refers only to samples submitted by the customer to SIRIM Berhad and tested by SIRIM Berhad. This report shall not be reproduced, except in full and shall not be used for advertising purposes by any means or forms without written approval from President & Chief Executive of SIRIM Berhad.




Figure 1 Relative Viability



MS ISO/IEC 17025
TESTING
SAMM NO.676

CONDITIONS RELATING TO THE USE OF SIRIM BERHAD TEST REPORT

1. A Test Report will be issued in respect of Testing Services conducted and shall related only to the sample actually tested. SIRIM Berhad makes no warranty whatsoever and the Applicant shall not represent in any manner that any duplication or mass production of the Product is same as the Sample actually tested or that SIRIM Berhad has tested any of the duplicated or mass produced Product.
2. The Test Report shall not be amended, changed, varied or modified in any manner whatsoever by the Applicant or otherwise.
3. If the Test Report is to be furnished to any third party or to the public, each such Test Report shall be furnished in full, legible and in its entirety.
4. The Test Report shall not be reproduced and shall not in any event be used for any advertising purposes or whatsoever without written approval from the President & Chief Executive of SIRIM Berhad of No. 1, Persiaran Dato' Menteri, Building 5, Section 2, P. O. Box 7035, 40700 Shah Alam, Selangor Darul Ehsan.
5. Customer (Applicant/Manufacturer/Factory, etc.) is not permitted to use any SIRIM Berhad, other SIRIM Berhad's subsidiaries logo on packaging, sample's manual, technical specification, brochures/flyers or any other means.
6. If such approval is obtained from the President & Chief Executive, the Applicant may only include the phrase, "A sample of this product has been tested by SIRIM Berhad ... (Test Report No) ... (dated) ... (for what test) ... (to which standard)" or such similar words which stress that only the Sample was actually tested. This phrase shall only be used for the purpose of product advertisement or product promotion (eg; brochures). For avoidance of doubt, the statement shall not be used on the sample and packaging of the sample.
7. In the event there is an investigation from a Government Regulatory Agency concerning the applicant's Test Report, SIRIM Berhad may disclose the information pertaining to the Test Report for purposes of such investigation.
8. Further or in the alternative, it is strictly forbidden to represent in any manner whatsoever that SIRIM Berhad and/or other SIRIM's subsidiaries has endorsed, approved or validated the Product of the Applicant in any manner whatsoever.
9. In the event the applicant is found in breach of this provision, SIRIM Berhad and/or other SIRIM's subsidiaries without prejudice to any other rights and remedies may take whatever action necessary including but not limited to:
 - a) Informing and placing a notice in the media;
 - b) Obtaining an injunction from Court (cost on a solicitor-client basis to be borne by the Applicant);
 - c) Refusing to accept any further Product for Testing Services from the Applicant or whatsoever related to the Applicant, whether subsidiary or otherwise;
 - d) Instructing the Applicant to withdraw and recall the advertisement, statement or document in question and advertise a clarification and apology to SIRIM Berhad and/or other SIRIM's subsidiaries twice in a national publication of SIRIM Berhad's choice at the Applicant's sole cost; and
 - e) Informing or lodging a report pertaining the Applicant's Test Report with the relevant authorities.
10. Certified true copies of the Test Report may be issued upon request by the applicant upon payment of the relevant fee.
11. Corrections to test report shall only be allowed within 6 months from issuance date of the Test Report of the relevant fee and shall be limited to maximum 3 times, after either case whichever occurs earlier, a new Test Report shall be issued and replace the previous one (having error(s) or lack of information) with relevant fee. Issuance of Supplementary Report to the original Test Report shall be for the followings:
 - a) Misprints and typo errors;
 - b) Missing technical information;
 - c) Test data not reported;
 - d) Mistake in reporting of test data.
12. Any amendment requested from customers on the test report issued shall be in writing.
13. SIRIM reserves the right in its sole discretion to terminate or modify this permission.