

HMK15 Humidity Calibrator



accepted and reliable method to calibrate humidity instruments. Usually two or three different salt solutions are used. Salts are chosen according to the application.

Certified Salts

The HMK15 can be ordered with certified and pre-measured salts. A sample calibration is made from each batch in Vaisala's Measurement Standards Laboratory (MSL).

Calibrated Thermometer

The HMK15 can be ordered with a thermometer, which is used for measuring the temperature during the calibration. It can also be used for checking the temperature measurement accuracy of the transmitter.

FINAS Accredited Measurement Standards Laboratory

Vaisala's Measurement Standards Laboratory is a FINAS accredited calibration laboratory. FINAS is a member of the EA (the European Cooperation for Accreditation).

Features/Benefits

- Easy and reliable calibration of humidity probes and transmitters
- Based on saturated salt solutions
- Fast temperature equilibration
- No external power required
- Suitable for laboratory use and on-site checks
- Chambers and transit covers make HMK15 easy to transport
- Pre-measured certified salts available
- Vaisala Service Centers offer accredited calibrations for humidity, temperature and barometric pressure.

No measuring instrument stays accurate by itself. It is essential that the functioning of an instrument is periodically checked against a reference. Vaisala has developed the Vaisala Humidity Calibrator HMK15 to make calibration and spot checking of humidity probes and transmitters easy and reliable.

Method Used by Leading Laboratories

The operating principle of the HMK15 is based on the fact that a saturated salt solution generates a certain relative humidity in the air above it. The reading of the humidity probe or transmitter can then be adjusted accordingly. Many leading laboratories use this generally



Technical Data

General

The standard HMK15 consists of the following parts:

- Two salt chambers, chamber covers and transit covers
- Base plate
- Thermometer
- Measurement cup and mixing spoon

Options

Certified and ready dosed salts:		Order code:	Total uncertainty*:
LiCl salt	11 %RH	19729HM	(±1.3 %RH)
MgCl ₂ salt	33 %RH	19730HM	(±1.2 %RH)
NaCl salt	75 %RH	19731HM	(±1.5 %RH)
K ₂ SO ₄ salt	97 %RH	19732HM	(±2.0 %RH)
*Uncertainties given at +20 °C			
Ion exchanged water		19767HM	
Extra salt chambers		19766HM	
Carrying bag		HM27032	
Set of rubber plugs		19746HM	

VAISALA

www.vaisala.com

Please contact us at
www.vaisala.com/requestinfo



Scan the code for
more information

Ref. B210908EN-D ©Vaisala 2016
This material is subject to copyright protection, with all
copyrights retained by Vaisala and its individual partners. All
rights reserved. Any logos and/or product names are trademarks
of Vaisala or its individual partners. The reproduction, transfer,
distribution or storage of information contained in this brochure
in any form without the prior written consent of Vaisala is strictly
prohibited. All specifications — technical included — are subject
to change without notice.

