

lifetime limited warranty

Procedure For Obtaining Warranty Services

1. In the event of a defect covered by this warranty, we will, at our option, repair or replace the device. This warranty does not cover device failure due to owner misuse or negligence, or normal wear and tear. The warranty does not extend to non-durable components, such as rubber accessories, casters, and grips, which are subject to normal wear and need periodic replacement. If you have a question about your Drive device or this warranty, please contact an authorized Drive dealer.
2. The complete (entire) Walker must be presented (returned) for inspection and evaluation. If the product is found to have defective materials or workmanship, the remedy will be limited to replacement or repair of the part needed to correct the problem. There will be no charge for the replaced or repaired materials.
3. Walkers which have been serviced under this warranty protection, will be returned to the owner at no additional cost for freight or delivery.
4. Warranty does not include any labor charges which may be incurred in the replacement or repair of parts or their installation. Freight or delivery charges to the location where such warranty services will be performed will be at the owners expense.
5. All Winnie Walker products are identified by a Lot Number at the time of manufacture. Removal or defacing the Lot Number may void the warranty.

For Service, in or out of warranty, complete the following information now for ready reference in the event correspondence becomes necessary. It is also required that you attach a copy of the Bill of Sale, should you be requested to show proof of purchase.

Dealers Name: _____
Address: _____
Phone No.: _____
Walker Lot No.: _____

lifetime limited warranty

Warranty Period

1. Your Drive brand product has lifetime protection against defective materials and workmanship covering the frame of the walker.

One (1) Year protection against defective materials or workmanship on all other components, assemblies and parts.

Limitations and Exclusions

Products which have been subject to negligence, abuse, improper storage or handling, improper operation, unauthorized modifications or damages beyond the reasonable control of Drive Products, are not covered by this warranty.

Parts or materials which are subject to normal wear resulting from the use of these products and which must be replaced/ repaired due to such normal wear, are the owners responsibility and are not covered by this Warranty.

This Warranty is exclusive and in lieu of all other express warranties, implied warranty, including but not limited to the implied warranty of merchantability and fitness for a particular purpose, and shall not extend beyond the duration of the express warranty provided herein. Drive Medical Design & Manufacturing shall not be liable for any consequential or incidental damages whatsoever.

This Warranty gives you specific legal rights which may vary from state to state. Some states do not allow the exclusions or limitations of incidental or consequential damages, or how long an implied warranty lasts, so the exclusions and limitation may not apply to you.



© 2006 Medical Depot, Inc. All rights reserved.
Drive is a trademark of Medical Depot, Inc.
Port Washington N.Y. 11050 USA Made in China

winnie lite supreme
aluminum rollator

drive
MEDICAL DESIGN & MANUFACTURING

www.drivemedical.com

mimi lite deluxe

The Mimi Lite Deluxe provides offers lightweight advantages and the added security of a locking loop brake. The easy folding and lightweight features combine to make this a great travel Rollator.

parts

- (2) handlebar
- (2) height adjustment knobs with bolts
- (1) basket.

assembly instructions

Please read these instructions before assembly or use of this Rollator. This Rollator must always be assembled and fitted for the user by a qualified Drive Medical Authorized dealer. Before using your new Rollator please ensure that your weight is within the weight capacity of this unit. This Rollator will require very little assembly and the assembly can be completed without tools.

1. Open your Rollator. Push the folding hinge until straight to secure the unit in the proper position.
2. Place a handle bar with the pre-assembled loop Lock in each handle bar receptacle and adjust to the desired height.
3. Use the height adjustment knobs and bolts to secure the handle bars in this position being cautious to ensure that the bolt heads are properly mounted within the hex opening on the inside of each handle bar receptacle.
4. Test and adjust the Loop Lock mechanism tension to ensure proper and secure Locking before use.
5. If your Rollator comes with a basket, mount the basket in the proper location.



Item # 199
Shown

folding

This unit is easily folded from the user's side of the unit by grasping the folding strap (black cloth strap) located under the hinged seat and lifting the unit by this strap. The unit will fold as lifting for easy storage and transport. The basket also is easily removed by a firm tug to pull the basket free of its frame clamps.

loop lock

This new generation loop lock handle has been designed to offer maximum comfort and safety. Two functions are incorporated: a normal loop lock and a parking loop lock. Those functions are fully separated, to avoid misunderstandings. Contact areas are wide, to limit pressure on the fingers.

Normal loop lock function may be applied, while walking with the Rollator, e.g. as a help to make a corner, or as a help to get up or off the side walk. Activating the Locks is performed by pulling the loop lock handle up, similar to operating the brakes of a bicycle.

The parking loop lock will be applied at standstill. When the user wants to sit down on the seat, the parking loop lock must be applied first, for safety reasons. If the user has to reach away from the Rollator, e.g. for taking something from the shelf in a shop, it will be safe to apply the parking loop lock first

The way of activating the parking loop lock is by pushing the handle down. This way is convenient for people who cannot easily pinch the handle, because of certain disabilities of their hands. Releasing the parking loop lock can be done by moving the handle up. For operation the parking loop lock, the loop lock handle is provided with a wider section at the end.

operation

The Rollator loop lock handle has been designed for operating the loop locks of a Rollator or walker and incorporates two functions: operating of normal loop lock and operating the parking loop lock. Normal loop lock will be used during walking, e.g. as help to make a corner. The parking loop lock function will be used at a standstill, e.g. to keep the Rollator at its place, when using the seat to sit down.


Characteristics of the loop lock handle are:

- 1) Normal loop lock function
- 2) Parking loop lock function
- 3) Operation of the normal loop lock and operation of parking loop lock are fully separated, to avoid misunderstandings.

Normal loop lock is activated by pulling the handle up, similar to operating the brakes of a bicycle. Once the handle is released, the lock will release as well.

-Activate the parking loop lock by pushing the handle DOWN. At the bottom position - the handle will lock automatically and the parking loop lock is ON.

-Release the parking loop lock by pulling the handle UP. The handle will return to its basic position; the parking loop lock is OFF.



NORMAL LOOP LOCK



PARKING LOOP LOCK

maintenance

Keep the loop lock handle free from dirt etc. to keep it operating properly. In case the loop locks are locking too little, the cable might be too loose. The cable can be adjusted by means of the adjuster screw, located where the cable is connected to the loop lock handle. If the loop lock cable wire is broken, it has to be replaced.

cautions

CAUTION: Don't use the loop lock handle for other functions than locking a Rollator or walker. Take care that the loop lock handle is in good condition, fully assembled (no loose parts) and fixed, for safety reasons.

Please read all cautions and warning labels before using this Rollator.

This unit is designed to be used as a mobility aid, creating a natural gate as the user walks securely with their primary weight distribution between the handlebars with the unit in front of the user.

If the Rollator has a seat, always engage the locking mechanism on the loop lock before using the seat. At no time can this unit be used to transport individuals while seated for risk of bodily injury.