

NATIONAL TYPE EVALUATION PROGRAM

Certificate of Conformance for Weighing and Measuring Devices

For:

Non-Computing Scale Digital Electronic

Model: Adventurer AXxNy n_{max}: 5200 to 220 000 (see below)

e_{min}: (see below)

Capacity: 220 g to 8200 g (see below)

Platform Size (see below) Accuracy Class: I, II

Submitted By:

Ohaus Corporation 7 Campus Drive Parsippany, NJ 07054

Tel: 973-377-9000 x 7032 Fax: 973-944-7177

Contact: Robert Hansen Email: bob.hansen@ohaus.com Web site: www.ohaus.com

Standard Features and Options

- Automatic Zero Tracking (AZT)
- Initial Zero Setting Mechanism (IZSM)
- Semi-Automatic Zero (Push Button)
- Semi-Automatic Tare (Push Button)
- Programmable Tare
- AC/DC Power Supply
- Gross/Net Display
- Automatic or Semi-Automatic Calibration
- Bracketing of the Display is Used to Identify "d" when it is not equal to "e" (d<e)
- Weight Units: carat, grain, gram, kilogram, milligram, pennyweight, pound, ounce, troy ounce.
- Model Description: E = External Calibration Weights Required, GN = Grain Test Scale

- Integral Display
- Liquid Crystal Display
- Separate Gross/Tare/Net Display
- Touch Screen Display
- **RS 232/USB**
- Linearity Calibration points (3)
- Weight Accumulation (Manual)
- "The Counting Feature is Not Legal for Trade" or "Counting Feature for Prescription Filling Only" is labeled on the front of the scale.
- Ethernet (optional)

Model	Capacity (g)	e	d	n _{max}	Class	Platter	Load Cell Used
						Dimensions	
AX224N	220 g	0.001 g	0.0001 g	220 000	I	90 mm	NMBA
AX423N	420 g	0.01 g	0.001 g	42 000	II	130 mm	AX523PC
AX223N/E	220 g	0.01 g	0.001 g	22 000	II	130 mm	AX523PE
AX423N/E	420 g	0.01 g	0.001 g	42 000	II	130 mm	AX523PE
AX523N/E	520 g	0.01 g	0.001 g	52 000	II	130 mm	AX523PE
AX522N/E	520 g	0.1 g	0.01 g	5200	II	175 x 195 mm	AX5202PE
AX622N/E	620 g	0.1 g	0.01 g	6200	II	175 x 195 mm	AX5202PE
AX822N/E	820 g	0.1 g	0.01 g	8200	II	175 x 195 mm	AX5202PE
AX1502N/E	1520 g	0.1 g	0.01 g	15 200	II	175 x 195 mm	AX5202PE
AX2202N/E	2200 g	0.1 g	0.01 g	22 000	II	175 x 195 mm	AX5202PE
AX4202N/E	4200 g	0.1 g	0.01 g	42 000	II	175 x 195 mm	AX5202PE
AX4202N/EGN	4200 g	0.1 g	0.01 g	42 000	II	175 x 195 mm	AX5202PE
AX8201N/E	8200 g	1 g	0.1 g	8200	II	175 x 195 mm	LSVK

Load Cells Used: Mettler Toledo non-NTEP (see table above) Temperature Range: 10 °C to 30 °C (50 °F to 86 °F)

This device was evaluated under the National Type Evaluation Program and was found to comply with the applicable technical requirements of "NIST Handbook 44: Specifications, Tolerances and Other Technical Requirements for Weighing and Measuring Devices." Evaluation results and device characteristics necessary for inspection and use in commerce are on the following pages.

Ronald Hayes


Chairman, NCWM, Inc.

Committee Chair, National Type Evaluation Program Committee

Issued: September 23, 2014

1135 M Street, Suite 110 / Lincoln, Nebraska 68508

The National Conference on Weights and Measures (NCWM) does not approve, recommend or endorse any proprietary product or material, either as a single item or as a class or group. Results shall not be used in advertising or sales promotion to indicate explicit or implicit endorsement of the product or material by the NCWM.







Ohaus Corporation

Non-Computing Scale / Adventurer AXxNy

<u>Application</u>: For general purpose weighing, retail jewelry/precious metal weighing, prescription weighing, prescription counting, weighing of grain in commercial and UDSA/GIPSA applications.

<u>Identification</u>: G.S.1 information is placed on a pressure sensitive or tamper proof identification badge located on the side of the device.

<u>Sealing</u>: A category 1 seal is used. To seal the device, a wire security seal may be threaded through tabs in a sliding cover and the base housing at the rear of the balance. Alternately, a destructible sealing label may be affixed to the sliding cover and the base housing. When the sliding cover is sealed in the locked position, access to the calibration switch inside the housing is prevented. Remote calibration and configuration are also blocked when the balance is sealed.

<u>Test Conditions</u>: This device was submitted to and evaluated by Measurement Canada under the U.S. and Canadian MRA. The technical data was reviewed by the Maryland NTEP laboratory for compliance with Publication 14 and NIST Handbook 44 requirements. The emphasis of the evaluation was on device design, operation, performance, permanence and compliance with influence factor requirements. Several increasing/decreasing load and shift tests were performed. The scales were tested over a temperature range of 10 °C to 30 °C (50 °F to 86 °F). Tests were also conducted with a power supply of 100VAC and 130VAC.

Evaluated By: J. Rae (MC), E.A.Payne, Jr (MD)

Type Evaluation Criteria Used: NIST Handbook 44 Specifications, Tolerances, and Other Technical Requirements for Weighing and Measuring Devices, 2014 Edition. NCWM Publication 14 Measuring Devices, 2014 Edition.

Conclusion: The results of the evaluation and information provided by the manufacturer indicate the device complies with applicable requirements.

Information Reviewed By: J. Truex (NCWM)

Examples of Device:



