

OIML Member State
The Netherlands

Number R76/2006-NL1-14.20
Project number 14200075
Page 1 of 2

| | |
|--------------------------------------|--|
| Issuing authority | NMi Certin B.V. Person responsible: C. Oosterman |
| Applicant and Manufacturer | Ohaus Corporation 7 Campus Drive, Suite 310 Parsippany, NJ 07054 United States of America |
| Identification of the certified type | A Non-automatic weighing instrument Type : Adventurer AX series |
| Characteristics | See next page |

This Certificate attests the conformity of the above identified Type (represented by the sample(s) identified in the OIML Test Report) with the requirements of the following Recommendation of the International Organization of Legal Metrology (OIML):

OIML R 76 - Edition 2006 for accuracy class **I** and **II**

This Certificate relates only to the metrological and technical characteristics of the type of measuring instrument covered by the relevant OIML International Recommendation above-identified. This Certificate does not bestow any form of legal international approval.

Important note: Apart from the mention of the Certificate's reference number and the name of the OIML Member State in which the Certificate was issued, partial quotation of the Certificate and of the associated OIML Test Report(s) is not permitted, although either may be reproduced in full.

Issuing Authority **NMi Certin B.V., OIML Issuing Authority NL1**
11 September 2014



C. Oosterman
Head Certification Board

NMi Certin B.V.
Hugo de Grootplein 1
3314 EG Dordrecht
the Netherlands
T +31 78 6332332
certin@nmi.nl
www.nmi.nl

This document is issued under the provision that no liability is accepted and that the applicant shall indemnify third-party liability.

The notification of NMi Certin B.V. as Issuing Authority can be verified at www.oiml.org

Parties concerned can lodge objection against this decision, within six weeks after the date of submission, to the general manager of NMi (see www.nmi.nl).



The conformity was established by the results of tests and examinations provided in the associated OIML Test Reports:

- No. NMI-14200075-01 dated 29 July 2014 that includes 41 pages;
- No. NMI-14200075-02 dated 29 July 2014 that includes 43 pages;
- No. NMI-14200075-03 dated 29 July 2014 that includes 33 pages;
- No. NMI-14200075-04 dated 29 July 2014 that includes 44 pages.

Characteristics of the non-automatic weighing instrument:

| Accuracy class | (I) | | (II) | |
|-----------------------------------|--|--|------------------------------|-------------------------|
| Maximum capacity | Max \leq 320 g | | Max \leq 5200 g | Max \leq 8200 g |
| Verification scale interval | $e \geq 0,001$ g | | $0,1$ g $\geq e \geq 0,01$ g | |
| Maximum number of scale intervals | $n \leq 320000$ divisions | | $n \leq 52000$ divisions | $n \leq 8200$ divisions |
| Actual scale interval | $e = d$ or $e = 10 d$ | | | |
| Weighing range | Single interval | | | |
| Tare | $T \leq -Max$ | | | |
| Temperature range | $+10$ °C / $+30$ °C | | | |
| Power supply voltage | 12 V DC supplied by AC/DC plug-in power supply | | | |
| Software identification | Version number: 1.xx, xx = 00 to 99 | | | |