

BODY:PVC, SOLMED:3294:PLUG:PC, MAKROLON: 2658 BY BAYER

2) VENDOR SHALL CERTIFY FOR EACH SHIPMENT THAT:

A) EACH SHIPMENT IS TRACEABLE TO VENDOR'S LOTS AND VENDOR'S SUPPLIER(S) LOTS

3) FLASH: NOT TO EXCEED .5MM 🔺

GATE LOCATION NOT TO EXCEED 1MM



- 4) MATERIAL IS COMPATIBLE WITH ETO AND GAMMA STERILIZATION
- 5) ALL DIMENSIONS ARE FOR REFERENCE ONLY UNLESS OTHERWISE SPECIFIED
- 6) PACKAGED: BULK, NON STERILE, DOUBLE POLY BAGS
- 7) UNLESS OTHERWISE SPECIFIED, ALL DIMENSIONS IN MM. + INDICATES INSPECTION DIM.

REV	DESCRIPTION:				
D	Changes idicated by ▲ per ECO 1623 7.				
VITALMED, INCORPORATED					
LAKEVILLE, MA 02347					
TWIST OFF BAG PORT WITH PLUG					

DRW. NO				
20391		Scale: N/A		REV.D
Chkd	RM 7/22/14	Appd	RR 7/22/14	<u>. </u>