



pH7

500ml / 17.5 fl.oz

Please READ and follow the instructions provided with your meter.

Leave your probe in the solution for a couple of minutes prior to calibration to adjust to temperature.

Use a clean container for solution and discard used solution when calibration is complete. Never return used solution to bottle.

Probe condition and battery strength will affect meter readings.

Test and replace probes as required.

Calibrate your meter regularly for best results.

KEEP OUT OF REACH OF CHILDREN!

Technical Support

Toll Free: 1.866.491.0255

Local: 250.491.0255

FUTURE HARVEST
NUTRIENTS

725 Evans Court,
Kelowna, BC, Canada V1Y 7G9
futureharvest.com



 Made in Canada