

Certificate of Conformance

Part No:
Revision:
Description:
Lot Number:

Manufactured Date:
Expiration Date:
Irradiation Dose: 25.0 – 45.0 kGy

Sample

- Biological:** The fluid path of this single use fluid transfer system has passed USP Class VI (USP<88>) and/or ISO 10993 testing.
- TSE/BSE:** The fluid path of this single use fluid transfer system conforms to European guidance EMA/410/01.
- Endotoxin:** The fluid path of a representative single use system is routinely tested in periodic validations for the presence of endotoxin in accordance with current USP Bacterial Endotoxin Test (USP<85>). Aqueous extracts contained <0.25 EU/mL as determined by the Limulus Amebocyte Lysate test (LAL).
- Particulate:** The fluid path of a representative single use system is routinely tested in periodic validations in accordance with current USP Particulate Matter in Injections Light Obscuration Particulate Count Test (USP<788>).
- Sterility:** The fluid path of the single use system has been validated following ANSI/AAMI/ISO 11137 guidelines for VDmax²⁵ to provide a minimum Sterility Assurance Level (SAL) of 10⁻⁶ for an established irradiation dose.
- Inspection:** This lot has been 100% visually inspected for acceptance in accordance with product specifications and Repligen Quality Control procedures.
- Manufacturing:** Repligen is an ISO 9001 certified company and products are manufactured meeting Repligen's quality management system.
This product was assembled in an ISO Class 7 cleanroom.
- Storage:** Store in original packaging at ambient conditions protected from moisture, extreme temperatures, and light.

Quality Assurance

Issue Date

HPK-QA-FM-10683-02