

## QUALITY TEST REPORT

品名 Product Name	医用外科口罩 Surgical Face Mask with Earloops	产品号 Product Code	401102	规格 Specification	17.5x9.5cm
客户订单号 Customer's Contract No.	29301	批号 Lot No.	29301		
测试时间 Testing time	2020-10-18	测试标准 Testing Standard	YY0469-2011		

项目 ITEM	标准值 STANDARD VALUE	测试结果 RESULT
原料 Material	67%聚丙烯无纺布, 33%熔喷布	OK
颜色 Color	蓝色 Blue	OK
尺寸 Size	17.5+/-0.2 cm	OK
	9.5 +/-0.2 cm	OK
层数 Layer	3 Layer	OK
每层克重 Weight of each layer	20g+25g+25g	OK
外观 Outer Appearance	毛发或者杂物不应出现在产品中 Hair and sundries should not existed in product.	OK
舒适性 Comfortable	口罩佩戴舒适 The mask shall be comfortable.	OK
橡筋 Rubber	橡筋位于口罩四个角, 佩戴方便舒适 The rubber shall be provided at each end of the mask for comfort and easy wearing.	OK
鼻梁架 Nose Bridge	10.5 cm	OK
橡筋长度 Length of Rubber	18 +/-2 cm	OK
橡筋弹性 Elasticity of Rubber	155%	OK
细菌过滤效率 Bacterial Filter Effect	≥98%, according to EN14683	OK

根据以上测试结果, 可以得出结论: 产品符合客户技术文档的要求。

According to above testing value, the conclusion is the products are in conformity with requirement of customer's specification.