



ACCURATE/WESTERN
SCALE CO. LTD.

CERTIFICATE OF ACCURACY

Certificate Number: 1004
Customer: Strongarm Sport Inc.
#114, 7880 - 66 St SE
Calgary, AB T2C 5V2
Contact: Mike Armstrong
Contact Number: 403-402-4142
Date Received: Nov 17/22
Date Checked: Nov 18/22
Date of Expiry:

Description of Weights: *Weightlifting*

Serial Number of Weight:

Condition of Weights: *new*

Make of scale used: *Excell*

Model of scale used: *CCT-10*

Serial of scale used: *202208006*

Wts used to calibrate scale: *AWS W.K. # 005*

Hereby certify the above scale has been calibrated to manufacturer's specifications.

Calibration is traceable to Measurement Canada Local Standards certified test standard

Serial #	Test Wt	Actual Wt	Error
<i>1.25 kg</i>	<i>1.25 kg</i>	<i>1.259 kg</i>	<i>+0.009 kg</i>
<i>2.5 kg</i>	<i>2.50 kg</i>	<i>2.510 kg</i>	<i>+0.010 kg</i>
<i>5 kg</i>	<i>5.00 kg</i>	<i>5.008 kg</i>	<i>+0.008 kg</i>
<i>10 kg</i>	<i>10.00 kg</i>	<i>10.011 kg</i>	<i>+0.011 kg</i>
<i>15 kg</i>	<i>15.00 kg</i>	<i>15.013 kg</i>	<i>+0.013 kg</i>
<i>20 kg</i>	<i>20.00 kg</i>	<i>20.005 kg</i>	<i>+0.005 kg</i>
<i>25 kg</i>	<i>25.00 kg</i>	<i>25.010 kg</i>	<i>+0.010 kg</i>