

	IFU-04-002 Rev. A 2023-07	
	<p>Instructions for Use <b><i>Read before use</i></b></p> <p>ACC-00-12500-00 30mm Female Tube Clamp</p>	
	PROTEOR SAS 6 rue de la Redoute 21850 Saint-Apollinaire – France +33 3 80 78 42 42 cs@proteor.com – www.proteor.com	
	PROTEOR USA 3 Morgan Irvine, CA 92618 – USA +1.855.450.7300 support@proteorusa.com – www.proteorusa.com	


	IFU-04-002 Rev. A 2023-07	
	<p>Instructions for Use <b><i>Read before use</i></b></p> <p>ACC-00-12500-00 30mm Female Tube Clamp</p>	
	PROTEOR SAS 6 rue de la Redoute 21850 Saint-Apollinaire – France +33 3 80 78 42 42 cs@proteor.com – www.proteor.com	
	PROTEOR USA 3 Morgan Irvine, CA 92618 – USA +1.855.450.7300 support@proteorusa.com – www.proteorusa.com	

	IFU-04-002 Rev. A 2023-07	
	<p>Instructions for Use <b><i>Read before use</i></b></p> <p>ACC-00-12500-00 30mm Female Tube Clamp</p>	
	PROTEOR SAS 6 rue de la Redoute 21850 Saint-Apollinaire – France +33 3 80 78 42 42 cs@proteor.com – www.proteor.com	
	PROTEOR USA 3 Morgan Irvine, CA 92618 – USA +1.855.450.7300 support@proteorusa.com – www.proteorusa.com	

	IFU-04-002 Rev. A 2023-07	
	<p>Instructions for Use <b><i>Read before use</i></b></p> <p>ACC-00-12500-00 30mm Female Tube Clamp</p>	
	PROTEOR SAS 6 rue de la Redoute 21850 Saint-Apollinaire – France +33 3 80 78 42 42 cs@proteor.com – www.proteor.com	
	PROTEOR USA 3 Morgan Irvine, CA 92618 – USA +1.855.450.7300 support@proteorusa.com – www.proteorusa.com	

**INTENDED USE/INDICATIONS**

The ACC-00-12500-00, 30mm female tube clamp is supplied to healthcare professionals (certified prosthetists/orthotists) to be used exclusively for the fitting of prosthetic lower limbs.

 This device is for multiple use on a single patient.

**COMPATIBILITIES**

The ACC-00-12500-00, 30mm female tube clamp is to be used with any standard 30mm pylon systems in conjunction with standard prosthetic componentry.

**ASSEMBLY**

Do not scratch the surface finish or otherwise alter the connector as these alterations may damage or weaken the device. Loctite 242 thread locker must be used on all screw threads during final assembly, and screws torqued appropriately.

Weight Limit	136 kg	(300 lb)
Alignment Set Screw Torque	15 Nm	(133 in-lb)
Pylon Clamp Cap Screw	7 Nm	(62 in-lb)

**⚠ WARNINGS: The clamp must be installed on the pyramid connector with the bolt in the posterior position. Failure to do so will potentially result in failures and injuries.**

**Any serious incident that occurs which relates to the device must be reported to the manufacturer and to the competent authority of the member state in which the incident occurred.**



**WARRANTY**

Proteor’s warranty of 24 months applies only if the product is used according to the specified condition and intended purpose, following all manufacturer’s recommendations. The products are tested according to ISO 10328.

**REGULATORY INFORMATION**


This product is a CE-marked medical device and is compliant with Regulation (EU) 2017/745.

**DESCRIPTION OF SYMBOLS**

	Manufacturer		Identified risk		CE marking
	Authorized representative in the European Union		Single patient, Multiple use		

**INTENDED USE/INDICATIONS**

The ACC-00-12500-00, 30mm female tube clamp is supplied to healthcare professionals (certified prosthetists/orthotists) to be used exclusively for the fitting of prosthetic lower limbs.

 This device is for multiple use on a single patient.

**COMPATIBILITIES**

The ACC-00-12500-00, 30mm female tube clamp is to be used with any standard 30mm pylon systems in conjunction with standard prosthetic componentry.

**ASSEMBLY**

Do not scratch the surface finish or otherwise alter the connector as these alterations may damage or weaken the device. Loctite 242 thread locker must be used on all screw threads during final assembly, and screws torqued appropriately.

Weight Limit	136 kg	(300 lb)
Alignment Set Screw Torque	15 Nm	(133 in-lb)
Pylon Clamp Cap Screw	7 Nm	(62 in-lb)

**⚠ WARNINGS: The clamp must be installed on the pyramid connector with the bolt in the posterior position. Failure to do so will potentially result in failures and injuries.**

**Any serious incident that occurs which relates to the device must be reported to the manufacturer and to the competent authority of the member state in which the incident occurred.**





**WARRANTY**

Proteor’s warranty of 24 months applies only if the product is used according to the specified condition and intended purpose, following all manufacturer’s recommendations. The products are tested according to ISO 10328.

**REGULATORY INFORMATION**


This product is a CE-marked medical device and is compliant with Regulation (EU) 2017/745.

**DESCRIPTION OF SYMBOLS**

	Manufacturer		Identified risk		CE marking
	Authorized representative in the European Union		Single patient, Multiple use		

**INTENDED USE/INDICATIONS**

The ACC-00-12500-00, 30mm female tube clamp is supplied to healthcare professionals (certified prosthetists/orthotists) to be used exclusively for the fitting of prosthetic lower limbs.

 This device is for multiple use on a single patient.

**COMPATIBILITIES**

The ACC-00-12500-00, 30mm female tube clamp is to be used with any standard 30mm pylon systems in conjunction with standard prosthetic componentry.

**ASSEMBLY**

Do not scratch the surface finish or otherwise alter the connector as these alterations may damage or weaken the device. Loctite 242 thread locker must be used on all screw threads during final assembly, and screws torqued appropriately.

Weight Limit	136 Kg	(300 lb)
Alignment Set Screw Torque	15 Nm	(133 in-lb)
Pylon Clamp Cap Screw	7 Nm	(62 in-lb)

**⚠ WARNINGS: The clamp must be installed on the pyramid connector with the bolt in the posterior position. Failure to do so will potentially result in failures and injuries.**

**Any serious incident that occurs which relates to the device must be reported to the manufacturer and to the competent authority of the member state in which the incident occurred.**






**WARRANTY**

Proteor’s warranty of 24 months applies only if the product is used according to the specified condition and intended purpose, following all manufacturer’s recommendations. The products are tested according to ISO 10328.

**REGULATORY INFORMATION**


This product is a CE-marked medical device and is compliant with Regulation (EU) 2017/745.

**DESCRIPTION OF SYMBOLS**

	Manufacturer		Identified risk		CE marking
	Authorized representative in the European Union		Single patient, Multiple use		

**INTENDED USE/INDICATIONS**

The ACC-00-12500-00, 30mm female tube clamp is supplied to healthcare professionals (certified prosthetists/orthotists) to be used exclusively for the fitting of prosthetic lower limbs.

 This device is for multiple use on a single patient.

**COMPATIBILITIES**

The ACC-00-12500-00, 30mm female tube clamp is to be used with any standard 30mm pylon systems in conjunction with standard prosthetic componentry.

**ASSEMBLY**

Do not scratch the surface finish or otherwise alter the connector as these alterations may damage or weaken the device. Loctite 242 thread locker must be used on all screw threads during final assembly, and screws torqued appropriately.

Weight Limit	136 Kg	(300 lb)
Alignment Set Screw Torque	15 Nm	(133 in-lb)
Pylon Clamp Cap Screw	7 Nm	(62 in-lb)

**⚠ WARNINGS: The clamp must be installed on the pyramid connector with the bolt in the posterior position. Failure to do so will potentially result in failures and injuries.**

**Any serious incident that occurs which relates to the device must be reported to the manufacturer and to the competent authority of the member state in which the incident occurred.**






**WARRANTY**

Proteor’s warranty of 24 months applies only if the product is used according to the specified condition and intended purpose, following all manufacturer’s recommendations. The products are tested according to ISO 10328.

**REGULATORY INFORMATION**

This product is a CE-marked medical device and is compliant with Regulation (EU) 2017/745.

**DESCRIPTION OF SYMBOLS**

	Manufacturer		Identified risk		CE marking
	Authorized representative in the European Union		Single patient, Multiple use		