

IFU-04-010 Rev. A 2023-07





IFU-04-010 Rev. A 2023-07





Instructions for Use *Read before use*

ACC-00-13800-00 3-Prong Rotating Pyramid Base



Instructions for Use Read before use

ACC-00-13800-00 3-Prong Rotating Pyramid Base



PROTEOR SAS 6 rue de la Redoute 21850 Saint-Apollinaire – France +33 3 80 78 42 42 cs@proteor.com – www.proteor.com



PROTEOR USA
3 Morgan
Irvine, CA 92618 – USA
+1.855.450.7300
support@proteorusa.com – www.proteorusa.com

EC REP

PROTEOR SAS 6 rue de la Redoute 21850 Saint-Apollinaire – France +33 3 80 78 42 42 cs@proteor.com – www.proteor.com



PROTEOR USA 3 Morgan Irvine, CA 92618 – USA +1.855.450.7300

support@proteorusa.com-www.proteorusa.com



IFU-04-010 Rev. A 2023-07





IFU-04-010 Rev. A 2023-07





Instructions for Use **Read before use**

ACC-00-13800-00 3-Prong Rotating Pyramid Base



Instructions for Use *Read before use*

ACC-00-13800-00 3-Prong Rotating Pyramid Base



PROTEOR SAS 6 rue de la Redoute 21850 Saint-Apollinaire – France +33 3 80 78 42 42 cs@proteor.com – www.proteor.com



PROTEOR USA
3 Morgan
Irvine, CA 92618 – USA
+1.855.450.7300
support@proteorusa.com – www.proteorusa.com

EC REP

PROTEOR SAS 6 rue de la Redoute 21850 Saint-Apollinaire – France +33 3 80 78 42 42 cs@proteor.com – www.proteor.com



PROTEOR USA
3 Morgan
Irvine, CA 92618 – USA
+1.855.450.7300
support@proteorusa.com – www.proteorusa.com

INTENDED USE/INDICATIONS

The ACC-00-13800-00, 3-Prong Rotating Pyramid Base is supplied to healthcare professionals (certified prosthetists/orthotists) to be used exclusively for the fitting of prosthetic lower limbs.



This device is for multiple use on a single patient.

COMPATIBILITIES

The ACC-00-13800-00, 3-Prong Rotating Pyramid Base is to be used with any standard 30mm or 34mm pylon systems in conjunction with standard prosthetic componentry.

ASSEMBLY

Do not scratch the surface finish or otherwise alter the connector as these alterations may damage or weaken the device. Loctite 242 thread locker must be used on all screw threads during final assembly, and screws torqued appropriately.

| Weight Limit | 136 kg | (300 lb) |
|-----------------------------|--------|------------|
| Rotation Clamp Screw Torque | 10 Nm | (88 in-lb) |

MARNINGS: The clamp must be installed on the pyramid connector with the bolt in the posterior position. Failure to do so will potentially result in failures and injuries.

Any serious incident that occurs which relates to the device must be reported to the manufacturer and to the competent authority of the member state in which the incident occurred.

WARRANTY

Proteor's warranty of 24 months applies only if the product is used according to the specified condition and intended purpose, following all manufacturer's recommendations. The products are tested according to ISO 10328.

REGULATORY INFORMATION

This product is a CE-marked medical device and is compliant with Regulation (EU) 2017/745.

DESCRIPTION OF SYMBOLS

| <u> </u> | Manufacturer | \triangle | Identified risk | ϵ | CE marking |
|----------|--|-------------|---------------------------------|------------|------------|
| EC REP | Authorized representative in the European Union | 1 | Single patient, Multiple use | | |

INTENDED USE/INDICATIONS

The ACC-00-13800-00, 3-Prong Rotating Pyramid Base is supplied to healthcare professionals (certified prosthetists/orthotists) to be used exclusively for the fitting of prosthetic lower limbs.



This device is for multiple use on a single patient.

COMPATIBILITIES

The ACC-00-13800-00, 3-Prong Rotating Pyramid Base is to be used with any standard 30mm or 34mm pylon systems in conjunction with standard prosthetic componentry.

ASSEMBLY

Do not scratch the surface finish or otherwise alter the connector as these alterations may damage or weaken the device. Loctite 242 thread locker must be used on all screw threads during final assembly, and screws torqued appropriately.

| Weight Limit Rotation Clamp Screw Torque | 136 kg | (300 lb) |
|---|--------|------------|
| Rotation Clamp Screw Torque | 10 Nm | (88 in-lb) |

 $ilde{\Delta}$ WARNINGS: The clamp must be installed on the pyramid connector with the bolt in the posterior position. Failure to do so will potentially result in failures and injuries.

Any serious incident that occurs which relates to the device must be reported to the manufacturer and to the competent authority of the member state in which the incident occurred.

WARRANTY

Proteor's warranty of 24 months applies only if the product is used according to the specified condition and intended purpose, following all manufacturer's recommendations. The products are tested according to ISO 10328.

REGULATORY INFORMATION

This product is a CE-marked medical device and is compliant with Regulation (EU) 2017/745.

DESCRIPTION OF SYMBOLS

| DECCION TION | | | | | |
|--------------|--|-------------|---------------------------------|------------|------------|
| F | Manufacturer | \triangle | Identified risk | ϵ | CE marking |
| EC REP | Authorized representative in the European Union | (11) | Single patient, Multiple use | | _ |

INTENDED USE/INDICATIONS

The ACC-00-13800-00, 3-Prong Rotating Pyramid Base is supplied to healthcare professionals (certified prosthetists/orthotists) to be used exclusively for the fitting of prosthetic lower limbs.



This device is for multiple use on a single patient.

COMPATIBILITIES

The ACC-00-13800-00, 3-Prong Rotating Pyramid Base is to be used with any standard 30mm or 34mm pylon systems in conjunction with standard prosthetic componentry.

ASSEMBLY

Do not scratch the surface finish or otherwise alter the connector as these alterations may damage or weaken the device. Loctite 242 thread locker must be used on all screw threads during final assembly, and screws torqued appropriately.

| Weight Limit | 136 kg | (300 lb) |
|-----------------------------|--------|------------|
| Rotation Clamp Screw Torque | 10 Nm | (88 in-lb) |

MARNINGS: The clamp must be installed on the pyramid connector with the bolt in the posterior position. Failure to do so will potentially result in failures and injuries.

Any serious incident that occurs which relates to the device must be reported to the manufacturer and to the competent authority of the member state in which the incident occurred.

WARRANTY

Proteor's warranty of 24 months applies only if the product is used according to the specified condition and intended purpose, following all manufacturer's recommendations. The products are tested according to ISO 10328.

REGULATORY INFORMATION

This product is a CE-marked medical device and is compliant with Regulation (EU) 2017/745.

DESCRIPTION OF SYMBOLS

| DESCRIPTION | OI STIVIDOLS | | | | |
|-------------|--|----------|---------------------------------|------------|------------|
| | Manufacturer | <u> </u> | Identified risk | ϵ | CE marking |
| EC REP | Authorized representative in the European Union | (11) | Single patient, Multiple use | | |

INTENDED USE/INDICATIONS

The ACC-00-13800-00, 3-Prong Rotating Pyramid Base is supplied to healthcare professionals (certified prosthetists/orthotists) to be used exclusively for the fitting of prosthetic lower limbs.



This device is for multiple use on a single patient.

COMPATIBILITIES

The ACC-00-13800-00, 3-Prong Rotating Pyramid Base is to be used with any standard 30mm or 34mm pylon systems in conjunction with standard prosthetic componentry.

ASSEMBL\

Do not scratch the surface finish or otherwise alter the connector as these alterations may damage or weaken the device. Loctite 242 thread locker must be used on all screw threads during final assembly, and screws torqued appropriately.

| Weight Limit | 136 kg | (300 lb) |
|-----------------------------|--------|------------|
| Rotation Clamp Screw Torque | 10 Nm | (88 in-lb) |

MARNINGS: The clamp must be installed on the pyramid connector with the bolt in the posterior position. Failure to do so will potentially result in failures and injuries.

Any serious incident that occurs which relates to the device must be reported to the manufacturer and to the competent authority of the member state in which the incident occurred.

WARRANTY

Proteor's warranty of 24 months applies only if the product is used according to the specified condition and intended purpose, following all manufacturer's recommendations. The products are tested according to ISO 10328.

REGULATORY INFORMATION

This product is a CE-marked medical device and is compliant with Regulation (EU) 2017/745.

DESCRIPTION OF SYMBOLS

| DESCRIT HOLE | OT STIVIDOES | | | | |
|--------------|--|-------------|---------------------------------|------------|------------|
| ш | Manufacturer | \triangle | Identified risk | ϵ | CE marking |
| EC REP | Authorized representative in the European Union | (1) | Single patient, Multiple use | | |
| | | | | | |