



# CERTIFICATE OF ACCREDITATION

## ANSI National Accreditation Board

11617 Coldwater Road, Fort Wayne, IN 46845 USA

This is to certify that

**iButtonLink, LLC**  
**1221 Innovation Drive, Suite 117**  
**Whitewater, WI 53190**

has been assessed by ANAB and meets the requirements of international standard

## ISO/IEC 17025:2017

while demonstrating technical competence in the field of

## CALIBRATION

Refer to the accompanying Scope of Accreditation for information regarding the types of activities to which this accreditation applies

AC-2574

Certificate Number

  
ANAB Approval

Certificate Valid Through: 04/03/2022  
Version No. 003 Issued: 02/28/2020



This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025:2017. This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (refer to joint ISO-ILAC-IAF Communiqué dated April 2017).



SCOPE OF ACCREDITATION TO ISO/IEC 17025:2017

iButtonLink, LLC
1221 Innovation Drive, Suite 117
Whitewater, WI 53190
Rob Olson (262) 662-4029
rolson@ibuttonlink.com

CALIBRATION

Valid to: April 3, 2022

Certificate Number: AC-2574

Thermodynamic

Table with 4 columns: Parameter/Equipment, Range, Expanded Uncertainty of Measurement (+/-), Reference Standard, Method, and/or Equipment. Row 1: Thermodynamic / Silicon or RTD Sensors, (-40 to 85) °C, 0.022 °C, Fluke 5628 SPRT & Readout.

Calibration and Measurement Capability (CMC) is expressed in terms of the measurement parameter, measurement range, expanded uncertainty of measurement and reference standard, method, and/or equipment. The expanded uncertainty of measurement is expressed as the standard uncertainty of the measurement multiplied by a coverage factor of 2 (k=2), corresponding to a confidence level of approximately 95%.

Notes:

- 1. This scope is formatted as part of a single document including Certificate of Accreditation No. AC-2574.

Handwritten signature of R. Olson
Vice President

