

## Certificate of Analysis 2015 Test Data

Method: EPA 200.8

Matrix: Water

Contaminant	% Reduction	MRL	Units	Dil	Analyzed	Qualifier
Antimony	99.7%	0.50	ug/l	1	10-1-15	CN-1
Arsenic	99.8%	0.40	ug/l	1	10-1-15	CN-1
Barium	99.9%	0.50	ug/l	1	10-1-15	CN-1
Cadmium	99.9%	0.10	ug/l	1	10-1-15	CN-1
Chromium	99.8%	0.20	ug/l	1	10-1-15	CN-1
Cobalt	99.9%	0.10	ug/l	1	10-1-15	CN-1
Copper	99.7%	0.50	ug/l	1	10-1-15	CN-1
Lead	99.9%	0.20	ug/l	1	10-1-15	CN-1
Molybdenum	99.9%	0.10	ug/l	1	10-1-15	CN-1
Nickel	99.6%	0.8	ug/l	1	10-1-15	CN-1
Selenium	99.7%	0.40	ug/l	1	10-1-15	CN-1
Thallium	99.9%	0.20	ug/l	1	10-1-15	CN-1
Vanadium	99.7%	0.50	ug/l	1	10-1-15	CN-1
Mercury	97.8%	0.50	ug/l	1	10-1-15	CN-1
Chromium 6+	99.4%	1.0	ug/l	1	10-1-15	CN-1
Nitrate	97.6%	110	ug/l	1	10-1-15	CN-1
Nitrite	92.0%	150	ug/l	1	10-1-15	CN-1
Fluoride, Total	89.36%	0.10	ug/l	1	10-1-15	CN-1
Benzene	98.7%	0.50	ug/l	1	12-22-15	CN-1
Chloroform	98.7%	0.50	ug/l	1	12-22-15	CN-1
TTHMs	99.0%	2.0	ug/l	1	12-22-15	CN-1



### Accreditations:

**NELAC #4047-002** (National Environmental Laboratory Accreditation Conference)

**ORELAP** (Oregon Environmental Laboratory Accreditation Program)

**ELAP #1132** (Environmental Laboratory Accreditation Program - A program managed by the State of California, Department of Health Services for accreditation of environmental testing laboratories)

**NEVADA #CA211**

**LACSD #10143** (County Sanitation Districts of Los Angeles County)