



SparMED Aps  
 Toppevadvej 34-38  
 DK-3660  
 Stenlose, Denmark



140 Hale Street  
 Haverhill, MA 01830  
[qc@embryotech.com](mailto:qc@embryotech.com)

ELI Accession Number: S2105-0315SPAR

Date of completion: 03-20-2015

Lot number: 03/2018

Reference number: OODW-70

Description of test article: Oosafe® Disinfection Wipes

**Assay system requested by customer:** An incubator was cleaned with test article. Post cleaning a prepared sperm sample was placed in the cleaned incubator containing one of the test articles for 24-hours.

**Test Assay method and results:** A non-frozen donor sperm specimen was obtained and used for this assay. The sperm was prepared and the motile fraction separated using the sperm gradient in a centrifuge cycle. Motility was noted at the beginning of the assay and again at 24-hours using a measure depth Makler Chamber System. Analyses were performed in sequence each time, with no more than 5-minutes between the test and the control sample.


**Results:**

	<u>Initial motility</u>	<u>24hour motility</u>
Test Article (incubator ELI-158)	93 %	93 %
Control (incubator ELI-068)	93 %	93 %

**Summary of observations:** The motility of the sperm remained consistent in the incubator that was cleaned with the test article and the control sperm that was incubated in ELI-068. Both the test and control showed no sign of affecting motility during the course of the assay.

  
 signature  
 Study Director

03-20-2015  
 date

  
 signature  
 Quality Reviewer

3/20/2015  
 date