

SparMED Aps Ryttermarken 2 3520 Farum Denmark



ELI Accession Number: SPAR-6078-0117

Date of completion: 01-09-2017

Lot numbers: 60465, 80350, 70355

Reference numbers: OODIH, OODH, OODSF

Description of test articles:

Oosafe CO2 Disinfectant, Oosafe Hand Disinfectant, Oosafe Surface Disinfectant

Assay system requested by customer: An incubator was cleaned with the test article (Oosafe CO2 Disinfectant). Post cleaning 1mL of each test article (Oosafe Hand Disinfectant and Oosafe Surface Disinfectant) was placed in a corresponding Petri dish and placed in the cleaned incubator. A culture plate was set up and one-cell mouse embryos were cultured in the cleaned incubator with the test articles for 96-hours.

Control assay method and results: 15 one cell (B₆C₃F₁ X B₆D₂F₁) embryos were cultured in triplicate micro drops of culture medium in control incubator ELI-248:

15 / 15 (100 %)

1-cell to 2-cell within 24 hr

15 / 15 (100 %)

1-cell to expanded blastocyst within 96 hr

For a valid assay, <u>Embryotech™</u> requires at least 70% of one cell control embryos to develop to expanded blastocyst within 96-hours.

Test assay method and results: 21 one cell (B₆C₃F₁ X B₆D₂F₁) embryos were cultured in triplicate micro drops of culture medium while in an incubator ELI-182 containing the Petri dishes filled with the test articles:

21 / 21 (100 %)

1-cell to 2-cell within 24 hr

20 / 21 (95 %)

1-cell to expanded blastocyst within 96 hr

Summary of observations: All test and control embryos were selected randomly from a common pool of freshly collected embryos. 100 percent of the control embryos developed to the expanded blastocyst stage within 96-hours. 95 percent of the embryos cultured in an incubator previously cleaned and containing the test article developed to the expanded blastocyst stage within 96-hours.

Signature

Study Director

01-09-2017

Date

Signature Quality Reviewer

Date