
3 Bledsoe


Tendon Strap


APPLICATION INSTRUCTIONS


CP020406 Rev B 08/08


This device is offered for sale by or on the order of a physician or other qualified medical professional. This device is not intended for reuse on a second patient. This device is not intended to prevent injuries or reduce or eliminate the risk of injury.

















INDICATIONS: Tennis Elbow Band Indicated for Tennis Elbow, Carpal Tunnel Syndrome and other forms of forearm and wrist pain. Infrapatellar Strap indicated for Patella Tendonitis and Chondromalacia.

CONTRAINDICATIONS: Tennis Elbow Band contraindicated for unstable fractures of the elbow. Infrapatellar Strap contraindicated for unstable fractures or for fractures of the proximal femur or the distal tibia or fibula.

WARNINGS: Do not leave this device in the trunk of a car on a hot day or in any other place where the temperature may exceed 150°F (65°C) for any extended period of time. If any additional pain or symptoms occur while using this brace, discontinue use and seek medical attention.

WARRANTY: This device is warranted to be free from defects in material and workmanship for a period of 90 days. This warranty applies to devices that have not been modified or subjected to misuse, abuse or neglect. This device is expressly subject to Bledsoe's Product Line Limited Warranty. Copies of this limited warranty can be obtained from your medical professional, via the internet at www.BledsoeBrace.com, or by calling 1.888.BLEDSOE (253.3763).

MANUFACTURED BY: BLEDSOE BRACE SYSTEMS™

2601 Pinewood; Grand Prairie, Texas, 75051, USA Toll Free Tel: 1.888.BLEDSOE (253.3763) • 1.800.527.3666 Local: 972.647.0884 • Local Fax: 972.660.5495

International Fax: 972.606.0649 www.BledsoeBrace.com

EUROPEAN AUTHORIZED REPRESENATIVE



EMERGO EUROPE Molenstraat 15 2513 BH, The Hague The Netherlands Tel: (+31) 70 345 8570 Fax: (+31) 70 346 7299