



Medical Device Details

Sponsor: CARSL Consulting
P O Box 766
Hastings
New Zealand

WAND Reference: 180615-WAND-6QIMJW

Sponsors Own Reference: Precise Medical Supplies - Tulip GEMS - Liposuction system

GMDN: Liposuction system [36894]

Class: IIa

Intended Purpose: Intended for the removal and placement of subcutaneous fat and fluid from the body for body sculpting.

Manufacturer: Tulip Medical Products 4360 Morean Boulevard, Suite 100
San Diego
California
UNITED STATES OF AMERICA

ARTG ID: 230146

> The device is supplied sterile.

- > Sterilisation method: Ethylene oxide
- > The device is intended to be invasive.
- > The device is intended for single use.
- > The device is not an active device.
- > The device does not contain material or ingredients of microbial origin.
- > The device does not contain material or ingredients of recombinant origin.
- > The device does not contain material or ingredients manufactured or formulated using a genetically modified organism.
- > The device does not contain material or ingredients of human origin.
- > The device does not contain human blood or its components.
- > The device consists of: Single product
- > The device does not contain material or ingredients of animal origin.
- > The device is not medicated.
- > The device is not formulated.
- > The product does not contain a medicine that has consent for marketing in New Zealand.
- > The product does not contain a medical device which incorporates a medicine as an integral part and that has an action ancillary to the device.

Status: Active