

Manual Supplement

Manual Title:	789/787B Calibration	Supplement Issue:	2
Print Date:	September 2002	Issue Date:	2/19
Revision/Date:	2, 1/17	Page Count:	1

This supplement contains information necessary to ensure the accuracy of the above manual.

Change #1, 173, 633

On page 6, under **AC Current Measurement**, replace the **Range** with:

Range

45 Hz to 500 Hz

On page 6, under **DC Current Measurement**, replace the Note with:

Note: 440 mA continuous, 1 A 30 seconds maximum on, 5 minutes off

Change #2, 668

On page 14, replace step 11 with:

11. Select the dc current function on the multimeter and set it to autorange range.