

# Supplement

Manual Title:	i3000s/i2000s Flex Instruction Sheet		
Part Number:	2560447	Supplement Issue:	<b>2</b>
Print Date:	March 2006	Issue Date:	7/19
Revision/Date:	1, 4/17	Page Count:	1

---

---

This supplement contains information necessary to ensure the accuracy of the above manual.

## Change #1, 451, 598

Under the **Specifications, Electrical Characteristics** section, replace Operating temperature and Storage temperature with:

Operating temperature: -20 °C to +50 °C

Storage temperature: -20 °C to +60 °C (without batteries)

Under **Measuring Head Characteristics** remove the Operating temperature and Storage temperature.

## Change #2, 739

On the 4<sup>th</sup> page, replace the Position sensitivity and add the figure:

Position sensitivity  $\pm 2\%$  of range with cable  
 $>25\text{ mm}$  (1 inch) from the coupling  
 (Center conductor perpendicularly  
 inside flexible probe area. If this is  
 not possible, an additional  $\pm 2\%$  of  
 full scale can occur.)  
 Measurements  $\leq 25\text{ mm}$  (1 inch)  
 from coupling are  $>3\%$ .

