






0-1 NTU TESTING GUIDE

- » Calibrate meter with 1 NTU Standard.
- » Use clean turbidity tubes [0260].
- » Do not pour standards in tubes back into the bottle.
- » Select Options: Nephelometric, Default Units, and Average measurements.
- » Use wall or USB adapter. Averaging option uses more power.
- » Read manual before testing.
- » To change Setup refer to manual.
- » For the most accurate results, follow the Tips.

Use only Code 0260 tubes.




CALIBRATION

1.	Press  to turn meter on .
2.	Select Measure Menu .
3.	Select Turbidity – With Blank .
4.	Rinse a tube [0260] three times with 0 NTU Standard or turbidity-free water. Fill the tube to the line with 0 NTU Standard or turbidity-free water. Cap the tube. This is the BLANK . Tip: Use a clean, smudge-free, scratch-free tube. Do not use a tube or cap that was used for high turbidity standards.
5.	Wipe the tube thoroughly with a lint-free cloth. Tip: Surround the tube with a clean, lint-free cloth. Press the cloth around the tube. Rotate the tube three times in the cloth to assure that all areas of the tube have been wiped.
6.	Insert the tube into the chamber. Close lid. Scan Blank . Remove the tube. Tip: Align the index line on the tube with the index arrow on the meter. Tip: After scanning the blank, scan the blank again as a sample. It should read 0.00. If not, reblank the meter and scan the blank again. Repeat until it reads 0.00. A small negative number will be observed if the reading is slightly less than the reading used as the blank. This is expected due to minute variations between readings.
7.	Empty the tube. Rinse the same tube three times with the 1 NTU Standard . Fill the tube to the line with 1 NTU Standard. Cap the tube. Tip: For the most accurate results, the same tube should be used for the Blank, 1 NTU Standard and the Sample to eliminate error caused by tube to tube variation. Tip: Fill the tube slowly, pouring down the inside wall of the tube to avoid introducing bubbles.
8.	Wipe the tube thoroughly with a lint-free cloth.
9.	Insert the tube into the chamber. Close the lid. Scan Sample . Tip: Scan the Sample three times, removing the tube from the chamber after each scan. The readings should be consistent. Use the last consistent reading to calibrate the meter.
10.	Press  . Select Calibrate .
11.	Press  or  to change the turbidity reading on the display to read 1.00 .
12.	Press  to save calibration.
13.	Proceed to Analysis.

0-1 NTU TESTING GUIDE (continued)

ANALYSIS (following calibration procedure)

1.	Press  to turn meter on .
	Tip: Meter should have been calibrated with 1.0 NTU Standard.
2.	Select Measure Menu .
3.	Select Turbidity – With Blank .
4.	Rinse a tube [0260] three times with 0 NTU Standard or turbidity-free water. Fill the tube to the line with 0 NTU Standard or turbidity-free water. Cap the tube. This is the BLANK .
	Tip: Use a clean, smudge-free, scratch-free tube. Do not use a tube or cap that was used for high turbidity standards.
5.	Wipe the tube thoroughly with a lint-free cloth.
	Tip: Surround the tube with a clean, lint-free cloth. Press the cloth around the tube. Rotate the tube three times in the cloth to assure that all areas of the tube have been wiped.
6.	Insert the tube into the chamber. Close lid. Scan Blank . Remove the tube.
	Tip: Align the index line on the tube with the index arrow on the meter. Tip: After scanning the blank, scan the blank again as a sample. It should read 0.00. If not, reblank the meter and scan the blank again. Repeat until it reads 0.00. A small negative number will be observed if the reading is slightly less than the reading used as the blank. This is expected due to minute variations between readings.
7.	Empty the tube. Rinse the same tube three times with the Sample . Fill the tube to the line with Sample. Cap the tube.
	Tip: For the most accurate results, the same tube should be used for the Blank, 1 NTU Standard and the Sample to eliminate error caused by tube to tube variation. Tip: Fill the tube slowly, pouring down the inside wall of the tube to avoid introducing bubbles.
8.	Wipe the tube thoroughly with a lint-free cloth.
9.	Insert the tube into the chamber. Close the lid. Scan Sample .
10.	Record the result.



802 Washington Ave | Chestertown | Maryland | 21620 · USA
800-344-3100 | 410-778-3100 (Outside U.S.A.) | Fax: 410-778-6394 | www.lamotte.com