
# eCheck - elepho

# Deluxe Pulse Oximeter

## ABOUT PULSE OXIMETERS

This pulse oximeter uses patented technology, which utilizes revolutionary optical sensor technology to measure both blood oxygen saturation (SpO<sub>2</sub>%) levels and pulse rate. By placing the finger into the sleeve and one-touch of a button, this device will send both red and infrared lights into the skin to collect data and display your blood oxygen and pulse rate.

### NAME/FUNCTION OF EACH PART



#### Other Accessories Included:

1 "1/2 AA" Lithium Battery (included in the unit)



Storage Pouch Lanyard

Note: When changing the orientation of the screen, the pulse rate will be shown on the top near the blinking Pulse Rate Icon, and the blood oxygen level will be shown on the bottom of the LCD display.

# **DISPLAY EXPLANATIONS**

	<b>Sensor Off Indicator:</b> Appears when the finger has been removed from the finger sleeve.
	<b>Weak Battery Symbol:</b> Flashes when battery should be replaced.
•	Pulse Rate: Shows the pulse rate per minute.
SpO <sub>2</sub>	<b>SpO<sub>2</sub> Indicator:</b> Shows the blood oxygen saturation level value on a scale of 70 to 99%.
	Calculating: Device is calculating blood oxygen

# REPLACING BATTERY

- 1. The battery compartment is located on the top of the unit. Remove battery cover by pressing down and sliding the battery cover off. (fig. 1)
- 2. Install or replace the battery according to the polarity indicator located inside the compartment. Use only 3.6V "1/2 AA" sized Lithium batteries. (fig. 2)
- 3. Replace the battery cover by inserting it back into place. (fig. 3)

#### Replace the batteries if:

- The weak battery symbol appears on the display.
- Nothing appears on the display after the finger is inserted into the sleeve and the button is pressed.

Note: Batteries are hazardous waste. Do not dispose of them together with the household garbage.







# MEASUREMENT PROCEDURE

If you are using this pulse oximeter for the first time, please remove the protective film from the screen and remove the pull tab from the battery door.

- 1. Hold the pulse oximeter so that the picture of the finger nail on the finger sleeve is facing up.
- 2. Gently place any finger into the finger sleeve as shown in the diagram (fig. A). Slightly press on the sides of the finger sleeve to ease the finger inside.
- 3. Turn the hand over so that the LCD screen is now facing you and your finger is pointed up (fig. B).
- 4. Push the Multi-directional button 🚳 to initiate readinas (fia. C).
- 5. Continue to vertically hold the pulse oximeter to enable readings
- 6. To change the direction of the LCD screen display by 180°, press the Multidirectional button 🐼 for up to two (2) seconds.

#### Note:

- Signal strength should be greater than one Fig.C bar to indicate a proper reading
- In the event of weak signal strength, please switch fingers and/or hand and try again.
- For greatest accuracy, limit hand motion during measurement.
- When the finger is removed, the pulse oximeter will turn off after 10 seconds.



Fig. A



Fig.B






Fig.D

# IMPORTANT PRODUCT NOTICES AND SAFETY INSTRUCTIONS

- Although the readings from this device are for reference only, the Mayo Clinic (http://mayoclinic.com/health/hypoxemia/MY00219) suggests that normal readings are between 95 to 100% under most circumstances. Readings under 90% are considered low.
- This is a home healthcare product only and is solely intended for sports and recreational use. It is not intended to serve as a substitute for the advice of a physician or medical professional.
- The readings from this pulse oximeter are for reference only. Consult a healthcare professional for interpretation of your pulse oximeter measurements. Contact a physician if you have or suspect any medical problem. Do not change your medications without the advice of your physician or healthcare professional.
- The pulse display is not suitable for checking the frequency of
- Use the pulse oximeter only for its intended use.
- Please keep this device away from children. Small parts such as battery and battery door could be hazardous if swallowed.
- The lanyard could cause strangulation in conditions when twisted around the neck.
- This pulse oximeter should not be used where alarms are required when taking measurements.
- For greatest accuracy, wait 10 minutes after exercising, bathing, eating, drinking beverages with alcohol or caffeine, or smoking to measure
- Use only 3.6V "1/2 AA" Lithium battery for power supply.
- Do not use cellular phones or other radio frequency producing devices near the working device as it may interfere with proper operation of the pulse oximeter.
- Not for use by or on persons under the age of 18 years old.

## CARE. MAINTENANCE & CLEANING

- Use a cotton swab with isopropyl alcohol solution (70%) to clean the inside of the finger sleeve or the outside of the unit.
- **DO NOT** use any other liquid or sprays on the oximeter for cleaning purposes.
- Under normal usage conditions, this device does not require calibration or technical maintenance other than infrequent replacement of the battery.
- Leaky batteries can damage the unit. Remove the battery when the unit will not be used for a long time.
- This product should only be used on the fingers.
- DO NOT disassemble this device. If in need of repair, refer to the warranty section of this manual.
- DO NOT use the product if the finger sleeve changes color or is no longer able to secure on finger.
- DO NOT subject this device to extreme shocks (do not drop on
- DO NOT drop or insert any object into any opening.
- DO NOT immerse the unit in water as this will result in damage to
- DO NOT subject this device to extreme temperatures, humidity, moisture, or direct sunlight. Protect from dust.
- If the unit is stored near freezing, allow it to acclimate to room temperature before use.
- This monitor may not meet its performance specifications if stored or used outside of these temperature and humidity ranges:

Storage Environment Temperature	Operation Environment Temperature	
-4°F - 158°F	39°F - 107°F	
(-20°C - 70°C)	(4°C - 42°C)	
Humidity: Less than 95%	Humidity: Less than 95%	
non-condensing	non-condensina	

# **TROUBLESHOOTING**

If any abnormality arises during use, please check the following points:

SYMPTOMS	POSSIBLE CAUSES	CORRECTION
Pulse Rate or SpO <sub>2</sub> is not appearing.	Position of the finger is not correct.	Reposition the finger into the correct position and measure again.
The pulse oximeter will not power on.	Battery has run down.	Replace the battery with a single 3.6V "½ AA" Lithium battery.
Screen orientation is upside down.	Viewing angle needs to be adjusted.	Press the Multi-directional button & for up to 2 seconds and the screen will rotate 180°.

# PRODUCT SPECIFICATIONS

Effective Measuring Range				
Saturation	70 to 99%			
Pulse Rate	40 to 250 bpm			
Resolution	1 digit			
Accuracy				
Saturation (70%-99%)	±2% or ±2 digits			
Pulse Rate	±3% or ±3 digits			
Display				
Saturation	2 characters			
Pulse Rate	3 characters			
Signal Indications:	Flashing Heart Icon, Sensor 'Off' and Signal intensity strength, battery indication - near empty			
Power Requirements	3.6V "1/2 AA" size Lithium battery			
Battery life	up to 300 Hours			
Type BF Applied Part	Spot Measurement No alarms			