



**Report For:** Dent-X Canada  
311 Bowes Road Main Unit  
Vaughan, Ontario  
L4K 1J1  
Phone: 416 774 2476  
Email: angelina@dent-xcanada.com

**Laboratory #:** 877327B-21  
FINAL  
**Report Date:** December 24, 2021  
**Received Date:** December 10, 2021  
**Customer P.O. #:** 3655

**Attention:** Angelina Wu  
**Specimen:** #2: FN-N95-508 Respirator

### TEST REPORT

One specimen, consisting of respirators were submitted to CMTL for assessment of mechanical headstrap strength to evaluate acceptability with both Health Canada performance criteria for filtering face piece respirators (Date published: 2020-08-25, Date modified: 2021-02-02) and 42 CFR Part 84 Subpart K, Sections 174 and 172 respectively.

Testing had also been conducted for particulate filter efficiency, airflow resistance, fluid resistance and flammability with results reported below from laboratory #877327D-21.



This report is subject to the following terms and conditions: 1. This report relates only to the specimen provided and there is no representation or warranty that it applies to similar substances or materials or the bulk of which the specimen is a part. 2. The content of this report is for the information of the customer identified above only and it shall not be reprinted, published or disclosed to any other party except in full. Prior written consent from Cambridge Materials Testing Limited is required. 3. The name Cambridge Materials Testing Limited shall not be used in connection with the specimen reported on or any substance or materials similar to that specimen without the prior written consent of Cambridge Materials Testing Limited. 4. Neither Cambridge Materials Testing Limited nor any of its employees shall be responsible or held liable for any claims, loss or damages arising in consequence of reliance on this report or any default, error or omission in its preparation or the tests conducted. 5. Specimens are retained 6 months, test reports and test data are retained 7 years from date of final test report and then disposed of, unless instructed otherwise in writing. 6. When making a statement of conformity to a specification or standard the report will make the statement of conformity based on the absolute value of the test result. Test Report Template Revision August 20, 2019


Per Steph Brown  
Authorized By Stephen Brown  
Per Derek Wild  
Technician, Derek Wild



Requirement for Filtering Facepiece Respirators per both  
Health Canada National Standard Specifications for Respirators during COVID-19:  
Guidance for Canadian Manufacturers (Date published: 2020-08-25, Date modified: 2021-02-02)  
and 42 CFR Part 84 Subpart K, Sections 172 and 174

<b>Characteristic</b>	<b>Barrier</b>	<b>Summary Results</b>
Particulate Filter Efficiency (%)	≥95	<b>Pass*</b>
Airflow (Inhalation) Resistance, mmH <sub>2</sub> O (Pa)	≤35 (343)	<b>Pass*</b>
Airflow (Exhalation) Resistance, mmH <sub>2</sub> O (Pa)	≤25 (245)	<b>Pass*</b>
Fluid Resistance maximum pressure in kPa for pass result	21.3	<b>Pass*</b>
Flammability, Class	1	<b>Pass*</b>
Mechanical Headstrap Strength, Observations and Proof Load (Newtons)	≥20	<b>Pass</b>

\* Test results reported under Laboratory #877327D-21 on December 17, 2021.



**PARTICULATE FILTRATION EFFICIENCY\***

Twenty submitted specimens were evaluated for particulate filtration efficiency in accordance with TEB-APR-STP-0059 test procedure to evaluate acceptability with Health Canada and 42 CFR 84 Subpart K requirements for N95 respirators.

All twenty of the specimens were conditioned (C) within a CSZ environmental control chamber for 25±1 hour at a 85±5% relative humidity and 38°C ± 2.5°C, then tested within 10 hours of extraction from the chamber as indicated in NIOSH standard procedure TEB-APR-STP-0059.

The particulate filter efficiency was performed on a TSI 8130A automated filter tester, and challenged under unidirectional airflow at 85 L/min ± 4 L/min with an aerosol of sodium chloride (NaCl) particles. The particles were generated by an aerosol generator and neutralized to their Boltzmann equilibrium state. The particles were considered to have an average count median diameter of 0.075 ± 0.020 micrometers and a geometric standard deviation not exceeding 1.86.

**RESULTS**

Specimen #	Conditioned	Flow Rate	Initial Filter Resistance (mmH <sub>2</sub> O)	Maximum Allowable Leakage (%)	Initial Leakage (%)	Maximum Leakage (%)	Particulate Filtration Efficiency (%)	Requirement (≥95%)	
								Result	Overall Result
1	C	85	9.7	5.00	1.755	1.755	98.2	Pass	Pass
2	C	85	9.1	5.00	1.966	1.966	98.0	Pass	
3	C	85	9.1	5.00	2.144	2.144	97.9	Pass	
4	C	85	9.6	5.00	1.670	1.670	98.3	Pass	
5	C	85	12.6	5.00	2.160	2.160	97.8	Pass	
6	C	85	9.2	5.00	1.997	1.997	98.0	Pass	
7	C	85	9.8	5.00	1.855	1.855	98.1	Pass	
8	C	85	12.2	5.00	2.172	2.172	97.8	Pass	
9	C	85	10.6	5.00	1.821	1.821	98.2	Pass	
10	C	85	9.6	5.00	1.991	1.991	98.0	Pass	
11	C	85	10.3	5.00	1.774	1.774	98.2	Pass	
12	C	85	10.5	5.00	1.887	1.887	98.1	Pass	
13	C	85	9.5	5.00	1.933	1.933	98.1	Pass	
14	C	85	9.4	5.00	1.723	1.723	98.3	Pass	
15	C	85	10.5	5.00	2.106	2.106	97.9	Pass	
16	C	85	12.4	5.00	2.294	2.294	97.7	Pass	
17	C	85	10.0	5.00	2.197	2.197	97.8	Pass	
18	C	85	9.9	5.00	1.920	1.920	98.1	Pass	
19	C	85	9.5	5.00	1.984	1.984	98.0	Pass	
20	C	85	9.7	5.00	1.893	1.893	98.1	Pass	

Note: As per Health Canada and 42 CFR Part 84 Subpart K, section 174(i) the minimum efficiency for each of the 20 filters will be determined and recorded and must be equal to or greater than 95% filtration efficiency.

\* Test results reported under Laboratory #877327D-21 on December 17, 2021.



**AIRFLOW (INHALATION) RESISTANCE\***


Twenty submitted specimens were evaluated for airflow (inhalation) resistance based on TEB-APR-STP-0007 using a TSI 8130A automated filter tester considered by NIOSH to be an acceptable pressure drop measurement.

Tests were performed with the salt generator turned-off under no loading conditions. Using hot-melt glue the filtering facepiece respirators were sealed onto flat plates with joint for connection to the resistance apparatus for measurements of pressure drop.

**RESULTS**

Specimen #	Maximum Allowable Resistance (mmH <sub>2</sub> O) Inhalation	Actual Resistance (mmH <sub>2</sub> O) Inhalation	Requirement (≤35)	
			Result	Overall Result
1	35	8.7	Pass	Pass
2	35	8.8	Pass	
3	35	9.0	Pass	
4	35	10.0	Pass	
5	35	10.0	Pass	
6	35	10.3	Pass	
7	35	11.9	Pass	
8	35	9.4	Pass	
9	35	9.1	Pass	
10	35	10.2	Pass	
11	35	10.7	Pass	
12	35	10.3	Pass	
13	35	10.4	Pass	
14	35	10.2	Pass	
15	35	12.5	Pass	
16	35	11.4	Pass	
17	35	12.1	Pass	
18	35	10.3	Pass	
19	35	10.1	Pass	
20	35	10.8	Pass	

\* Test results reported under Laboratory #877327D-21 on December 17, 2021.



**AIRFLOW (EXHALATION) RESISTANCE\***


Twenty submitted specimens were evaluated for airflow (exhalation) resistance based on TEB-APR-STP-0003 using a TSI 8130A automated filter tester considered by NIOSH to be an acceptable pressure drop measurement.

Tests were performed with the salt generator turned-off under no loading conditions. Using hot-melt glue the filtering facepiece respirators were sealed onto flat plates, and mounted in reverse, with joint for connection to the resistance apparatus for measurements of pressure drop.

**RESULTS**

Specimen #	Maximum Allowable Resistance (mmH <sub>2</sub> O) Exhalation	Actual Resistance (mmH <sub>2</sub> O) Exhalation	Requirement (≤25)	
			Result	Overall Result
1	25	8.8	Pass	Pass
2	25	8.8	Pass	
3	25	9.0	Pass	
4	25	9.5	Pass	
5	25	9.3	Pass	
6	25	9.4	Pass	
7	25	9.5	Pass	
8	25	9.3	Pass	
9	25	9.0	Pass	
10	25	8.9	Pass	
11	25	9.1	Pass	
12	25	9.4	Pass	
13	25	9.4	Pass	
14	25	9.1	Pass	
15	25	9.8	Pass	
16	25	9.4	Pass	
17	25	9.7	Pass	
18	25	9.3	Pass	
19	25	9.4	Pass	
20	25	9.4	Pass	

\* Test results reported under Laboratory #877327D-21 on December 17, 2021.



**FLUID RESISTANCE**

ASTM F1862/F1862M-17 at 21.3-kPa (160mmHg) Pressure

**RESULTS**

Specimen #	Test Pressure (mmHg)	Total Number of Specimens	Number of Pass Specimens
4	160	32	32

Note: Acceptable Quality Limit of 4.0% is met for single sampling plan when 29 or more of the 32 tested specimens show pass results.

<b>Material construction type</b>	FN-N95-510H
<b>Supplier</b>	Dent-X Canada
<b>Lot number</b>	A10B7CW11/29/2021 1577
<b>Date of receipt</b>	December 10, 2021
<b>Date of test</b>	December 17, 2021
<b>Fluid velocity (cm/s)</b>	647
<b>Volume of impact fluid (ml)</b>	2
<b>Angle of pneumatic valve to horizontal</b>	3°
<b>Description target area mask</b>	Outer Surface
<b>Distance from tip cannula to mask (in)</b>	12
<b>Technique to enhance visual detection</b>	Cotton swab used to lightly daub on the surface
<b>Conditioning parameters</b>	21±5°C, 85±5% R.H for minimum of 4 hours

**NOTE:** The outside surface of the mask is exposed to the blood stream in order to observe whether penetration occurred on the inner surface of the mask that could be contacting the wearer's face. Penetration on the inner facing of the mask constitutes failure (ASTM F1862/F1862M-17 section 4.2).

\* Test results reported under Laboratory #877327D-21 on December 17, 2021.



**FLAMMABILITY**

The specimen, consisting of 5 masks, was tested in accordance to 16 CFR 1610 (1-1-16 Edition).

**RESULTS**


Specimen #4	Specimen #	RESULT	CONCLUSION
	4-1	IBE	Classified as Class 1
	4-2	IBE	
	4-3	IBE	
	4-4	IBE	
	4-5	IBE	

**IBE:** Ignited but extinguished

**Test:** Flame Resistance 45° angle test. One-Second Flame Impingement.  
**Type of fabric:** Without a raised fiber surface  
**Surface tested:** Face  
**Type of test:** Original State  
**Direction tested:** Length  
**Testing Conditioning:** Specimens conditioned at 105°C for 30 min, then placed in desiccator  
**Requirements:** The flame spread time for textile products without a raised fibre surface must be greater than 3.5 seconds.  
**Date of Receipt:** December 10, 2021  
**Date of Test:** December 17, 2021

Note: For a test plan of 5 specimens, no failure is allowed for an Acceptable Quality Limit of 4.0%.

\* Test results reported under Laboratory #877327D-21 on December 17, 2021.



**MECHANICAL HEADSTRAP STRENGTH**



<b>Report For:</b>	Cambridge Materials Testing Limited 6991 Millcreek Drive, Unit 13 MISSISSAUGA, Ontario L5N 6B9	<b>Laboratory #:</b>	879175-21
<b>Attention:</b>	Derek Wild	<b>Report Date:</b>	December 24, 2021
<b>Specimen:</b>	Respirators, CMTL Mississauga Lab # 877327, Customer: Dent-X Canada, Identified as #2-FN-N95-508 Respirator	<b>Received Date:</b>	December 15, 2021
		<b>Customer P.O.#:</b>	

**PROOF LOAD TEST REPORT**

Ten submitted specimens were subjected to proof load testing in accordance with Health Canada National Standard Specifications for Respirators during COVID-19: Guidance for Canadian Manufacturers, Date Published: February 2, 2021. Testing was performed by donning the mask body on to a head form. A proof load of 10 N was then applied to the elastomeric strap for 10 seconds. The proof load was then removed and the specimen was examined for failure. Testing machine was operated in accordance with ASTM A370-20 paragraph 8 with a test speed of 75mm/minute.

**RESULTS**

Specimen	Observations
1	There was no evidence of breakage, tearing, separation from the point of fixation to the respirator body, permanent deformation or other obvious loss of function in the securing mechanism.
2	There was no evidence of breakage, tearing, separation from the point of fixation to the respirator body, permanent deformation or other obvious loss of function in the securing mechanism.
3	There was no evidence of breakage, tearing, separation from the point of fixation to the respirator body, permanent deformation or other obvious loss of function in the securing mechanism.
4	There was no evidence of breakage, tearing, separation from the point of fixation to the respirator body, permanent deformation or other obvious loss of function in the securing mechanism.
5	There was no evidence of breakage, tearing, separation from the point of fixation to the respirator body, permanent deformation or other obvious loss of function in the securing mechanism.
6	There was no evidence of breakage, tearing, separation from the point of fixation to the respirator body, permanent deformation or other obvious loss of function in the securing mechanism.
7	There was no evidence of breakage, tearing, separation from the point of fixation to the respirator body, permanent deformation or other obvious loss of function in the securing mechanism.
8	There was no evidence of breakage, tearing, separation from the point of fixation to the respirator body, permanent deformation or other obvious loss of function in the securing mechanism.
9	There was no evidence of breakage, tearing, separation from the point of fixation to the respirator body, permanent deformation or other obvious loss of function in the securing mechanism.
10	There was no evidence of breakage, tearing, separation from the point of fixation to the respirator body, permanent deformation or other obvious loss of function in the securing mechanism.

**File Name**

This report is subject to the following terms and conditions: 1. This report relates only to the specimen provided and there is no representation or warranty that it applies to similar substances or materials or the bulk of which the specimen is a part. 2. The content of this report is for the information of the customer identified above only and it shall not be reprinted, published or disclosed to any other party except in full. Prior written consent from Cambridge Materials Testing Limited is required. 3. The name Cambridge Materials Testing Limited shall not be used in connection with the specimen reported on or any substance or materials similar to that specimen without the prior written consent of Cambridge Materials Testing Limited. 4. Neither Cambridge Materials Testing Limited nor any of its employees shall be responsible or held liable for any claims, loss or damages arising in consequence of reliance on this report or any default, error or omission in its preparation or the tests conducted. 5. Specimens are retained 6 months, test reports and test data are retained 7 years from date of final test report and then disposed of, unless instructed otherwise in writing. 6. When making a statement of conformity to a specification or standard the report will make the statement of conformity based on the absolute value of the test result. Test Report Template Revision: January 14, 2019

Cambridge Materials Testing Limited

Per Nicholas Wolfenberg Authorized By  
Nicholas Wolfenberg  
Per Brayden Dahmer Technician  
Brayden Dahmer