

www.gendepot.com

Certificate of Analysis

TBS Buffer (10X), pH 7.4

Cat No : T8057 Lot No : 01052413

TEST	Specification	Result
Appearance	Clear, colorless liquid	PASS
pH @ 25°C : 10X Solution	7.4 <u>+</u> 0.15	7.41
Concentration (1X)		
Tris	25 mM <u>+</u> 0.0025	PASS
Sodium Chloride	140 mM <u>+</u> 0.014	PASS

This document has been produced electronically and is valid without signature

Date of Print: 1.10.2024