

# Certificate of Analysis

## amfiSure ONE PCR Master Mix(2X)

Cat No : P7000

Lot No : 01192424

Exp. Feb. 2027

TEST	Specification	Result
Appearance	Light red liquid	PASS
PCR Performance	Single band	PASS
Tracking Dyes Migration	Red/Yellow	PASS
PCR / DNA Electrophoresis condition		
Reaction volume : 25 ul		
Template DNA Source : K562		
Primer : b-Globin		
Loading volume on the gel : 10ul		
Gel Staining : Safe-Pinky DNA gel staining , 30 min		
Target PCR Band Size : 536 bp.		
25mM NaOH Tolerance Test		PASS
PCR Performed with added 2ul of 2.5mM NaOH for 20ul reaction.		
GC-Rich DNA Sequence PCR		PASS
PCR Performed with 75% GC rich DNA sequence 3Kb , 50ul reaction.		

This document has been produced electronically and is valid without signature

Date of Print : 1. 23. 2024