

**Product Name : 10X PBS Buffer**

Cat No : P2100

Lot No : 03182412

Exp. Mar. 2027

TEST	Specification	Result
Appearance	Clear, Colorless Solution	PASS
Sterility		
Aerobic	Negative	NEGATIVE
Anaerobic	Negative	NEGATIVE
pH , 10X Solution <i>pH was taken at room temperature(~ 22°C)</i>	6.7 - 6.9	6.8
Osmolality	2800 - 3200 mM/kg	3010 mM/kg
Endotoxin <i>Endotoxin testing was performed using LAL detection method.</i>	≤ 1.0 EU/ml	< 0.1 EU/ml
Mycoplasma <i>Mycoplasma test was performed using MycoAlert® Mycoplasma Detection Assay</i>	Negative	NEGATIVE
Animal Cell Culture Toxicity Functionality	Negative ≥ 80%	NEGATIVE PASS

This document has been produced electronically and is valid without signature

Date of Print : 3. 21. 2024