

Certificate of Analysis

Product Name : Trypsin-EDTA(1X), 0.25%

www.gendepot.com

Cat No : CA014	Lot No : 03222418	Exp. Apr. 2026
TEST	Specifications	Results
Appearance	Clear, Red Solution	PASS
Sterility		
Aerobic	Negative	NEGATIVE
Anaerobic	Negative	NEGATIVE
pH pH was taken at room temperature(~ 22°C)	7.20 - 8.00	7.7
Trypsin Activity	<u>></u> 250 U/mg	648 U/mg
Chymotrypsin Activity	≥50 to≤120	72 U/mg
Osmolality	270 - 305 mM/kg	315 mM/kg
Endotoxin Endotoxin testing was performed using LAL detection method.	Record	< 0.1EU/ml
Mycoplasma Mycoplasma test was performed using MycoAlert [®] Mycoplasma Detection Assay	Negative	NEGATIVE
Animal Cell Culture Functionality	Detaches Cells with \ge 80% viability	PASS

This document has been produced electronically and is valid without signature

Date of Print : 3. 27. 2024