

Certificate of Analysis

Triton X-100

Cat No : T9500

Lot No : 03222421

TEST	Specification	Result
Appearance	Clear viscous liquid	PASS
CMC(H ₂ O)	0.01 –0.016 %	PASS
Aggregation Number (H ₂ O)	75 –165	PASS
Color (APHA)	≤ 100	8
Protease	None	PASS
Cloud Point	Report	66 °C
pH (5%, Water) @ 25°C	6.0 - 8.0	7.1

This document has been produced electronically and is valid without signature

Date of Print : 4. 2. 2024