

Product Name : Poly-D-Lysine Solution in PBS, 0.01%

Cat No : CA084

Lot No : 03242416

Exp. Apr. 2026

TEST	Specifications	Results
Appearance	Clear Solution	PASS
Sterility		
Aerobic	Negative	NEGATIVE
Anaerobic	Negative	NEGATIVE
pH <i>pH was taken at room temperature (~ 22°C)</i>	7.35 - 7.45	7.4
Osmolarity	280 - 310 mM/kg	298 mM/kg
Endotoxin <i>Endotoxin testing was performed using LAL detection method.</i>	< 0.1 EU/ml	≤ 0.1 EU/ml
PDI by GPC		1.035
MN by NMR		53100
DP _n by NMR		254
Mycoplasma		NEGATIVE

This document has been produced electronically and is valid without signature

Date of Print : 4. 3. 2024