

Certificate of Analysis

Mini Spin Column without lid

Cat No : S1910

Lot No : 04261911

TEST	Specification	Result
Plasmid DNA Extraction	30 - 50 ug	35 ug
DNA recovery from Agarose gel	80 - 95 %	88 %
DNA recovery from PCR reaction	85 - 97 %	93 %
DNase / RNase	None detected	None
Transparency	Opaque	PASS
Rigidity	14000 RPM 30 min	PASS

This document has been produced electronically and is valid without signature

Date of Print : 4. 29. 2019