

Certificate of Analysis

TBST(10X), 0.5% Tween 20, pH 7.4

Cat No : T8058

Lot No : 05262314

TEST	Specification	Result
Appearance	Clear, colorless liquid	PASS
pH @ 25°C : 10X Solution	7.4 ± 0.25	7.40
Concentration (1X)		
Tris	25 mM ± 0.0025	PASS
Sodium Chloride	140 mM ± 0.014	PASS
Tween 20	0.5%	PASS

This document has been produced electronically and is valid without signature

Date of Print : 5. 23. 2023